

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035783.D
 Acq On : 22 May 2023 14:38
 Operator : JC/MD
 Sample : 02860-05DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-10-131-05152023DL

Quant Time: May 23 06:12:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	211224	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	375516	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144678	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	183352	49.758	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.520%
35) Dibromofluoromethane	5.385	113	120375	47.533	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	8.646	98	435731	46.574	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.140%
62) 4-Bromofluorobenzene	11.079	95	171662	45.571	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.140%
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	2620	1.654	ug/l #	87
44) Trichloroethene	7.122	130	130744	46.109	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035783.D
Acq On : 22 May 2023 14:38
Operator : JC/MD
Sample : 02860-05DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-10-131-05152023DL

Quant Time: May 23 06:12:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

