

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035787.D
 Acq On : 22 May 2023 16:12
 Operator : JC/MD
 Sample : 02860-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EQB-11-05162023

Quant Time: May 23 06:13:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	170323	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	325926	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	168198	56.607	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.220%	
35) Dibromofluoromethane	5.385	113	110371	50.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.420%	
50) Toluene-d8	8.647	98	375678	46.265	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 92.540%	
62) 4-Bromofluorobenzene	11.079	95	149394	45.693	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.380%	
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	3470	2.717	ug/l	# 87
20) Methylene Chloride	2.794	84	1011	0.437	ug/l	# 82
29) Tetrahydrofuran	5.019	42	20401	18.360	ug/l	95
30) Chloroform	5.093	83	193105	44.288	ug/l	100
47) Bromodichloromethane	7.824	83	5556	1.634	ug/l	# 83
52) Toluene	8.726	92	1286	0.210	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035787.D
Acq On : 22 May 2023 16:12
Operator : JC/MD
Sample : 02860-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EQB-11-05162023

Quant Time: May 23 06:13:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

