

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035790.D
 Acq On : 22 May 2023 17:22
 Operator : JC/MD
 Sample : 02860-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-05-101-0512023

Quant Time: May 23 06:14:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	195192	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	359944	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	308271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	179533	52.723	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	105.440%	
35) Dibromofluoromethane	5.385	113	108911	44.866	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	89.740%	
50) Toluene-d8	8.647	98	414752	46.250	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	92.500%	
62) 4-Bromofluorobenzene	11.079	95	157947	43.744	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	87.480%	
Target Compounds						
					Qvalue	
16) Acetone	2.410	43	14277	9.755	ug/l	92
18) Methyl Acetate	2.715	43	1589	0.294	ug/l	91
25) 2-Butanone	4.587	43	33789	16.678	ug/l	95
44) Trichloroethene	7.129	130	4398	1.618	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035790.D
Acq On : 22 May 2023 17:22
Operator : JC/MD
Sample : 02860-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-05-101-0512023

Quant Time: May 23 06:14:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

