

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035791.D
 Acq On : 22 May 2023 17:45
 Operator : JC/MD
 Sample : 02860-11 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-DUP-03-05162023

Quant Time: May 23 06:14:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	159047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	301515	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	267523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	112799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156834	56.524	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.040%
35) Dibromofluoromethane	5.385	113	99251	48.810	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.620%
50) Toluene-d8	8.647	98	350265	46.628	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.260%
62) 4-Bromofluorobenzene	11.079	95	138467	45.780	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.560%
Target Compounds						
						Qvalue
16) Acetone	2.404	43	1548	1.298	ug/l	93
20) Methylene Chloride	2.788	84	3637	1.683	ug/l #	75
44) Trichloroethene	7.123	130	26268	11.538	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035791.D
Acq On : 22 May 2023 17:45
Operator : JC/MD
Sample : 02860-11 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-DUP-03-05162023

Quant Time: May 23 06:14:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

