

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035793.D
 Acq On : 22 May 2023 18:32
 Operator : JC/MD
 Sample : PB152922TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB152922TB

Quant Time: May 23 06:15:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	192060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	350339	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	322272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	175042	52.243	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.480%	
35) Dibromofluoromethane	5.385	113	113039	47.844	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.680%	
50) Toluene-d8	8.647	98	417143	47.792	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	95.580%	
62) 4-Bromofluorobenzene	11.079	95	171079	48.680	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	97.360%	
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	45573	31.647	ug/l	91
20) Methylene Chloride	2.788	84	11069	4.241	ug/l	94
37) Ethyl Acetate	4.721	43	54119	13.913	ug/l	99
43) Isopropyl Acetate	6.342	43	221418	32.643	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035793.D
Acq On : 22 May 2023 18:32
Operator : JC/MD
Sample : PB152922TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB152922TB

Quant Time: May 23 06:15:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

