

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035795.D
 Acq On : 22 May 2023 19:18
 Operator : JC/MD
 Sample : 02864-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-280

Quant Time: May 23 06:16:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	130722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	242014	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108644	47.641	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.280%
35) Dibromofluoromethane	5.385	113	77617	47.556	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.120%
50) Toluene-d8	8.646	98	280881	46.584	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.160%
62) 4-Bromofluorobenzene	11.079	95	101014	41.608	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.220%
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	27554	28.112	ug/l	99
20) Methylene Chloride	2.794	84	3575	2.013	ug/l #	80
37) Ethyl Acetate	4.714	43	14524	5.405	ug/l	99
43) Isopropyl Acetate	6.342	43	69027	14.731	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035795.D
Acq On : 22 May 2023 19:18
Operator : JC/MD
Sample : 02864-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-280

Quant Time: May 23 06:16:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

