

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
 Data File : VX035798.D
 Acq On : 22 May 2023 20:28
 Operator : JC/MD
 Sample : 02884-05 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2004

Quant Time: May 23 06:17:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	166240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	310593	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	280756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134411	46.347	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.700%		
35) Dibromofluoromethane	5.385	113	98610	47.078	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.160%		
50) Toluene-d8	8.647	98	359370	46.442	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.880%		
62) 4-Bromofluorobenzene	11.079	95	134870	43.287	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.580%		
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	5483	4.399	ug/l	89
20) Methylene Chloride	2.794	84	2105	0.932	ug/l #	83
40) Benzene	6.044	78	3702	0.400	ug/l #	88
52) Toluene	8.714	92	1556	0.267	ug/l	95
68) m/p-Xylenes	10.293	106	857	0.201	ug/l	97
95) Naphthalene	13.774	128	6088	0.653	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035798.D
Acq On : 22 May 2023 20:28
Operator : JC/MD
Sample : 02884-05 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2004

Quant Time: May 23 06:17:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

