



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052223\
Data File : VX035773.D
Acq On : 22 May 2023 10:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Title : SW846 8260
Last Update : Thu May 11 13:57:54 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	14592	PASS
75	95	30	60	54.9	34995	PASS
95	95	100	100	100.0	63688	PASS
96	95	5	9	6.6	4205	PASS
173	174	0.00	2	1.1	451	PASS
174	95	50	100	64.4	41043	PASS
175	174	5	9	7.3	2997	PASS
176	174	95	101	95.9	39363	PASS
177	176	5	9	6.0	2356	PASS