

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
 Data File : VX046313.D
 Acq On : 22 May 2025 08:51
 Operator : JC/MD
 Sample : VX0522WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0522WBL01

Quant Time: May 23 01:39:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	120286	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48946	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62210	53.923	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.840%
35) Dibromofluoromethane	5.385	113	44210	51.040	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.080%
50) Toluene-d8	8.647	98	155104	51.736	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.480%
62) 4-Bromofluorobenzene	11.079	95	57952	50.394	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
Data File : VX046313.D
Acq On : 22 May 2025 08:51
Operator : JC/MD
Sample : VX0522WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0522WBL01

Quant Time: May 23 01:39:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

