

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
 Data File : VX046320.D
 Acq On : 22 May 2025 11:37
 Operator : JC/MD
 Sample : Q2097-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR200044

Quant Time: May 23 01:43:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125758	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64415	54.059	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.120%	
35) Dibromofluoromethane	5.379	113	46228	51.047	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.100%	
50) Toluene-d8	8.647	98	161904	51.654	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.300%	
62) 4-Bromofluorobenzene	11.079	95	62473	51.961	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.920%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	39879	83.471	ug/l	96
18) Methyl Acetate	2.709	43	2252	2.031	ug/l	93
20) Methylene Chloride	2.788	84	2633	2.876	ug/l	90
43) Isopropyl Acetate	6.348	43	11737	5.118	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
Data File : VX046320.D
Acq On : 22 May 2025 11:37
Operator : JC/MD
Sample : Q2097-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR200044

Quant Time: May 23 01:43:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

