

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
 Data File : VX046322.D
 Acq On : 22 May 2025 12:23
 Operator : JC/MD
 Sample : Q2097-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-244

Quant Time: May 23 01:44:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63541	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125999	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62880	53.081	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.160%	
35) Dibromofluoromethane	5.385	113	46214	50.934	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.860%	
50) Toluene-d8	8.647	98	159674	50.845	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.700%	
62) 4-Bromofluorobenzene	11.079	95	64000	53.130	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	106.260%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	42097	88.630	ug/l	97
18) Methyl Acetate	2.709	43	2253	2.044	ug/l	91
20) Methylene Chloride	2.788	84	2652	2.914	ug/l #	74
43) Isopropyl Acetate	6.348	43	12371	5.384	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
Data File : VX046322.D
Acq On : 22 May 2025 12:23
Operator : JC/MD
Sample : Q2097-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-244

Quant Time: May 23 01:44:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

