

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
 Data File : VX046324.D
 Acq On : 22 May 2025 13:10
 Operator : JC/MD
 Sample : Q2097-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT3419

Quant Time: May 23 01:45:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	60281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	121358	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52335	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63603	56.595	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.180%
35) Dibromofluoromethane	5.379	113	45023	51.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.040%
50) Toluene-d8	8.647	98	157835	52.182	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.360%
62) 4-Bromofluorobenzene	11.079	95	62290	53.687	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.380%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	41735	92.620	ug/l	100
18) Methyl Acetate	2.703	43	2658	2.542	ug/l	95
20) Methylene Chloride	2.782	84	2411	2.792	ug/l	94
43) Isopropyl Acetate	6.342	43	12011	5.427	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
Data File : VX046324.D
Acq On : 22 May 2025 13:10
Operator : JC/MD
Sample : Q2097-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT3419

Quant Time: May 23 01:45:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

