

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
 Data File : VX046328.D
 Acq On : 22 May 2025 14:43
 Operator : JC/MD
 Sample : Q2101-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1-MHE

Quant Time: May 23 01:47:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	69289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138151	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69775	54.015	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.040%	
35) Dibromofluoromethane	5.385	113	50952	51.217	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.440%	
50) Toluene-d8	8.647	98	172796	50.184	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.360%	
62) 4-Bromofluorobenzene	11.079	95	65514	49.602	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.200%	
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	41799	80.703	ug/l	97
18) Methyl Acetate	2.703	43	2377	1.978	ug/l	96
20) Methylene Chloride	2.782	84	5497	5.538	ug/l	92
43) Isopropyl Acetate	6.348	43	12832	5.093	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052225\
Data File : VX046328.D
Acq On : 22 May 2025 14:43
Operator : JC/MD
Sample : Q2101-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1-MHE

Quant Time: May 23 01:47:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

