

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
 Data File : VX009853.D
 Acq On : 23 May 2019 13:10
 Operator : JC/SP
 Sample : VX0523MBS01
 Misc : 5.0µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01

Quant Time: May 24 03:49:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	184671	50.00	µg/l	0.00
34) 1,4-Difluorobenzene	6.86	114	298584	50.00	µg/l	0.00
63) Chlorobenzene-d5	10.11	117	266547	50.00	µg/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	129886	50.00	µg/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.06	65	116773	46.32	µg/l	0.00
Spiked Amount	50.000		Recovery	=	92.64%	
35) Dibromofluoromethane	5.50	113	82592	44.00	µg/l	0.00
Spiked Amount	50.000		Recovery	=	88.00%	
50) Toluene-d8	8.71	98	337759	44.57	µg/l	0.00
Spiked Amount	50.000		Recovery	=	89.14%	
62) 4-Bromofluorobenzene	11.13	95	119903	43.54	µg/l	0.00
Spiked Amount	50.000		Recovery	=	87.08%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.20	85	27523	16.173	µg/l	97
3) Chloromethane	1.32	50	42058	18.102	µg/l	99
4) Vinyl Chloride	1.41	62	40543	17.895	µg/l	98
5) Bromomethane	1.64	94	25627	18.846	µg/l	100
6) Chloroethane	1.72	64	25867	17.783	µg/l	100
7) Trichlorofluoromethane	1.93	101	49336	17.915	µg/l	99
8) Diethyl Ether	2.19	74	22923	17.573	µg/l	94
9) 1,1,2-Trichlorotrifluoroet	2.38	101	30041	17.371	µg/l	99
10) Methyl Iodide	2.51	142	32233	19.162	µg/l	98
11) Tert butyl alcohol	3.05	59	62260	106.693	µg/l	98
12) 1,1-Dichloroethene	2.37	96	31265	18.044	µg/l	94
13) Acrolein	2.29	56	40491	76.618	µg/l	98
14) Allyl chloride	2.73	41	77103	18.259	µg/l	98
15) Acrylonitrile	3.14	53	128663	90.175	µg/l	99
16) Acetone	2.45	43	123931	91.552	µg/l	98
17) Carbon Disulfide	2.57	76	87477	17.982	µg/l	99
18) Methyl Acetate	2.78	43	57295	17.797	µg/l	98
19) Methyl tert-butyl Ether	3.20	73	119357	17.739	µg/l	99
20) Methylene Chloride	2.85	84	37677	17.190	µg/l	97
21) trans-1,2-Dichloroethene	3.17	96	33369	17.757	µg/l	95
22) Diisopropyl ether	3.85	45	149689	18.433	µg/l	93
23) Vinyl Acetate	3.81	43	668874	96.187	µg/l	97
24) 1,1-Dichloroethane	3.70	63	70789	17.670	µg/l	99
25) 2-Butanone	4.67	43	197573	93.923	µg/l	97
26) 2,2-Dichloropropane	4.59	77	50785	17.031	µg/l	99
27) cis-1,2-Dichloroethene	4.60	96	39527	17.622	µg/l	99
28) Bromochloromethane	5.01	49	34006	19.410	µg/l	95
29) Tetrahydrofuran	5.13	42	128041	94.591	µg/l	98
30) Chloroform	5.21	83	65004	17.941	µg/l	100
31) Cyclohexane	5.58	56	69441	18.631	µg/l	94
32) 1,1,1-Trichloroethane	5.49	97	52843	17.520	µg/l	98
36) 1,1-Dichloropropene	5.80	75	49947	18.000	µg/l	100
37) Ethyl Acetate	4.84	43	72221	18.681	µg/l	99
38) Carbon Tetrachloride	5.79	117	43957	17.436	µg/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
 Data File : VX009853.D
 Acq On : 23 May 2019 13:10
 Operator : JC/SP
 Sample : VX0523MBS01
 Misc : 5.0µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01

Quant Time: May 24 03:49:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	59758	17.536	ug/l	94
40) Benzene	6.14	78	152947	17.810	ug/l	99
41) Methacrylonitrile	5.04	41	41006	17.818	ug/l	98
42) 1,2-Dichloroethane	6.20	62	56392	18.306	ug/l	99
43) Isopropyl Acetate	6.45	43	112719	18.227	ug/l	99
44) Trichloroethene	7.21	130	35389	17.046	ug/l	97
45) 1,2-Dichloropropane	7.51	63	43926	18.064	ug/l	98
46) Dibromomethane	7.66	93	24743	17.523	ug/l	100
47) Bromodichloromethane	7.90	83	48050	17.059	ug/l	98
48) Methyl methacrylate	7.77	41	57434	18.720	ug/l	97
49) 1,4-Dioxane	7.74	88	21944	352.687	ug/l	96
51) 4-Methyl-2-Pentanone	8.64	43	380605	94.162	ug/l	97
52) Toluene	8.78	92	91160	17.463	ug/l	99
53) t-1,3-Dichloropropene	9.04	75	55667	17.505	ug/l	99
54) cis-1,3-Dichloropropene	8.43	75	61183	17.382	ug/l	92
55) 1,1,2-Trichloroethane	9.21	97	37637	17.673	ug/l	98
56) Ethyl methacrylate	9.18	69	67312	17.706	ug/l	96
57) 1,3-Dichloropropane	9.37	76	65707	17.533	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.31	63	172208	85.391	ug/l	99
59) 2-Hexanone	9.49	43	293807	92.551	ug/l	98
60) Dibromochloromethane	9.58	129	35213	16.929	ug/l	97
61) 1,2-Dibromoethane	9.67	107	37585	17.484	ug/l	98
64) Tetrachloroethene	9.34	164	33062	18.083	ug/l	98
65) Chlorobenzene	10.14	112	93599	17.554	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.22	131	33591	17.377	ug/l	99
67) Ethyl Benzene	10.25	91	176949	18.032	ug/l	99
68) m/p-Xylenes	10.36	106	129305	36.058	ug/l	98
69) o-Xylene	10.69	106	62101	17.552	ug/l	97
70) Styrene	10.71	104	107775	17.526	ug/l	97
71) Bromoform	10.85	173	26673	16.979	ug/l #	100
73) Isopropylbenzene	11.02	105	169623	17.959	ug/l	100
74) N-amyl acetate	10.90	43	100318	18.218	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.27	83	62028	17.617	ug/l	99
76) 1,2,3-Trichloropropane	11.29	75	69914	23.205	ug/l	83
77) Bromobenzene	11.26	156	41226	17.585	ug/l	99
78) n-propylbenzene	11.36	91	192041	17.803	ug/l	100
79) 2-Chlorotoluene	11.42	91	115269	17.321	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	142619	17.933	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.07	75	19038	16.168	ug/l	92
82) 4-Chlorotoluene	11.51	91	133508	17.758	ug/l	99
83) tert-Butylbenzene	11.77	119	132756	17.160	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	144063	18.047	ug/l	99
85) sec-Butylbenzene	11.94	105	161747	17.635	ug/l	100
86) p-Isopropyltoluene	12.06	119	147548	17.894	ug/l	99
87) 1,3-Dichlorobenzene	12.02	146	73304	17.540	ug/l	100
88) 1,4-Dichlorobenzene	12.10	146	73963	17.554	ug/l	99
89) n-Butylbenzene	12.39	91	131831	17.689	ug/l	99
90) Hexachloroethane	12.60	117	23154	16.482	ug/l	96
91) 1,2-Dichlorobenzene	12.39	146	73430	17.251	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	13.00	75	14795	18.731	ug/l	93

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
Data File : VX009853.D
Acq On : 23 May 2019 13:10
Operator : JC/SP
Sample : VX0523MBS01
Misc : 5.0µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0523MBS01

Quant Time: May 24 03:49:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	51345	17.639	ug/l	99
94) Hexachlorobutadiene	13.78	225	25804	17.359	ug/l	98
95) Naphthalene	13.83	128	167033	17.638	ug/l	99
96) 1,2,3-Trichlorobenzene	14.02	180	52120	17.664	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
 Data File : VX009853.D
 Acq On : 23 May 2019 13:10
 Operator : JC/SP
 Sample : VX0523MBS01
 Misc : 5.0µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 Client Sample ID :
 VX0523MBS01

Quant Time: May 24 03:49:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration









#1
 Pentafluorobenzene
 Concen: 50.000 ug/l
 RT: 5.66 min Scan# 739
 Delta R.T. -0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleID :
 VX0523MBS01


Tgt Ion: 168 Resp: 184671
 Ion Ratio Lower Upper
 168 100
 99 56.8 42.3 63.5



#2
 Dichlorodifluoromethane
 Concen: 16.173 ug/l
 RT: 1.20 min Scan# 7
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 85 Resp: 27523
 Ion Ratio Lower Upper
 85 100
 87 30.5 16.1 48.3









#3
 Chloromethane
 Concen: 18.102 ug/l
 RT: 1.32 min Scan# 28
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion: 50 Resp: 42058
 Ion Ratio Lower Upper
 50 100
 52 31.6 25.9 38.9



#4
 Vinyl Chloride
 Concen: 17.895 ug/l
 RT: 1.41 min Scan# 42
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 62 Resp: 40543
 Ion Ratio Lower Upper
 62 100
 64 31.8 24.6 36.8









#5
 Bromomethane
 Concen: 18.846 ug/l
 RT: 1.64 min Scan# 79
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
ClientSampled :
 VX0523MBS01


Tgt Ion: 94 Resp: 25627
 Ion Ratio Lower Upper
 94 100
 96 91.0 73.0 109.6



#6
 Chloroethane
 Concen: 17.783 ug/l
 RT: 1.72 min Scan# 93
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 64 Resp: 25867
 Ion Ratio Lower Upper
 64 100
 66 31.4 25.3 37.9








#7
 Trichlorofluoromethane
 Concen: 17.915 ug/l
 RT: 1.93 min Scan# 127
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10





Instrument : MSVOA_X
 ClientSampleId : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
101	49336		
103	64.9	52.4	78.6



#8
 Diethyl Ether
 Concen: 17.573 ug/l
 RT: 2.19 min Scan# 170
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
74	22923		
45	112.9	53.3	159.9









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 17.371 ug/l
 RT: 2.38 min Scan# 201
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
101	30041		
85	44.9	35.7	53.5
151	78.6	62.2	93.2



#10
 Methyl Iodide
 Concen: 19.162 ug/l
 RT: 2.51 min Scan# 222
 Delta R.T. 0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
142	32233		
127	42.2	32.7	49.1
141	14.9	11.5	17.3









#11
 Tert butyl alcohol
 Concen: 106.693 ug/l
 RT: 3.05 min Scan# 311
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
59	100		
57	9.4	8.2	12.2



#12
 1,1-Dichloroethene
 Concen: 18.044 ug/l
 RT: 2.37 min Scan# 200
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
96	100		
61	169.1	144.3	216.5
98	61.5	50.3	75.5









#13
 Acrolein
 Concen: 76.618 ug/l
 RT: 2.29 min Scan# 187
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
56	40491		
55	72.5	56.6	85.0



#14
 Allyl chloride
 Concen: 18.259 ug/l
 RT: 2.73 min Scan# 258
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
41	77103		
39	59.2	46.7	70.1
76	25.7	21.8	32.8









#15
 Acrylonitrile
 Concen: 90.175 ug/l
 RT: 3.14 min Scan# 326
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
53	128663		
52	83.3	66.9	100.3
51	36.3	28.2	42.2



#16
 Acetone
 Concen: 91.552 ug/l
 RT: 2.45 min Scan# 212
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
43	123931		
58	30.0	24.9	37.3









#17
 Carbon Disulfide
 Concen: 17.982 ug/l
 RT: 2.57 min Scan# 232
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampleID : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
76	100		
78	9.7	7.4	11.0



#18
 Methyl Acetate
 Concen: 17.797 ug/l
 RT: 2.78 min Scan# 266
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
43	100		
74	19.4	16.2	24.4









#19
 Methyl tert-butyl Ether
 Concen: 17.739 ug/l
 RT: 3.20 min Scan# 335
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
73	119357		
57	23.7	19.3	28.9



#20
 Methylene Chloride
 Concen: 17.190 ug/l
 RT: 2.85 min Scan# 279
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
84	37677		
49	142.9	109.9	164.9
51	42.2	32.7	49.1
86	64.2	52.0	78.0









#21
 trans-1,2-Dichloroethene
 Concen: 17.757 ug/l
 RT: 3.17 min Scan# 330
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
96	33369		
Ion Ratio			
96	100		
61	163.2	124.5	186.7
98	65.0	50.4	75.6



#22
 Diisopropyl ether
 Concen: 18.433 ug/l
 RT: 3.85 min Scan# 443
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
45	149689		
Ion Ratio			
45	100		
43	78.1	56.1	84.1
87	22.7	18.1	27.1
59	10.3	8.5	12.7









#23
 Vinyl Acetate
 Concen: 96.187 ug/l
 RT: 3.81 min Scan# 436
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
43	100		
86	7.6	6.9	10.3



#24
 1,1-Dichloroethane
 Concen: 17.670 ug/l
 RT: 3.70 min Scan# 417
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
63	100		
98	6.7	3.0	9.2
100	3.7	2.0	6.0









#25
 2-Butanone
 Concen: 93.923 ug/l
 RT: 4.67 min Scan# 577
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
43	197573		
72	21.3	18.2	27.2



#26
 2,2-Dichloropropane
 Concen: 17.031 ug/l
 RT: 4.59 min Scan# 563
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
77	50785		
97	22.6	11.5	34.4









#27
 cis-1,2-Dichloroethene
 Concen: 17.622 ug/l
 RT: 4.60 min Scan# 565
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
96	39527		
Ion Ratio			
96	100		
61	161.9	0.0	324.0
98	63.1	0.0	130.4



#28
 Bromochloromethane
 Concen: 19.410 ug/l
 RT: 5.01 min Scan# 633
 Delta R.T. 0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
49	34006		
Ion Ratio			
49	100		
129	1.7	0.0	3.8
130	60.1	51.2	76.8









#29
 Tetrahydrofuran
 Concen: 94.591 ug/l
 RT: 5.13 min Scan# 652
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
42	128041		
72	37.1	30.4	45.6
71	33.9	28.6	43.0



#30
 Chloroform
 Concen: 17.941 ug/l
 RT: 5.21 min Scan# 666
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
83	65004		
85	64.2	51.5	77.3









#31
 Cyclohexane
 Concen: 18.631 ug/l
 RT: 5.58 min Scan# 726
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
56	100		
69	28.6	23.4	35.0
84	70.3	61.9	92.9



#32
 1,1,1-Trichloroethane
 Concen: 17.520 ug/l
 RT: 5.49 min Scan# 712
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
97	100		
99	64.4	51.8	77.8
61	49.7	37.8	56.8









#33
 1,2-Dichloroethane-d4
 Concen: 46.321 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
65	116773		
67	52.8	0.0	106.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
114	298584		
63	22.4	0.0	45.6
88	16.5	0.0	33.2









#35
 Dibromofluoromethane
 Concen: 44.002 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
113	82592		
111	102.4	82.8	124.2
192	21.4	17.0	25.6



#36
 1,1-Dichloropropene
 Concen: 18.000 ug/l
 RT: 5.80 min Scan# 762
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
75	49947		
110	33.7	16.9	50.6
77	30.8	24.8	37.2









#37
 Ethyl Acetate
 Concen: 18.681 ug/l
 RT: 4.84 min Scan# 604
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
43	72221		
61	13.5	10.5	15.7
70	9.2	7.9	11.9



#38
 Carbon Tetrachloride
 Concen: 17.436 ug/l
 RT: 5.79 min Scan# 760
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
117	43957		
119	98.2	77.5	116.3
121	30.8	24.7	37.1









#39
 Methylcyclohexane
 Concen: 17.536 ug/l
 RT: 7.46 min Scan# 1035
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
83	59758		
83	100		
55	91.3	67.7	101.5
98	42.9	35.5	53.3



#40
 Benzene
 Concen: 17.810 ug/l
 RT: 6.14 min Scan# 818
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
78	152947		
78	100		
77	23.1	18.8	28.2









#41
 Methacrylonitrile
 Concen: 17.818 ug/l
 RT: 5.04 min Scan# 638
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
41	100		
39	54.7	42.0	63.0
67	62.0	49.8	74.8
52	27.9	20.3	30.5



#42
 1,2-Dichloroethane
 Concen: 18.306 ug/l
 RT: 6.20 min Scan# 827
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
62	100		
98	8.8	0.0	18.2









#43
 Isopropyl Acetate
 Concen: 18.227 ug/l
 RT: 6.45 min Scan# 868
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
43	112719		
61	18.6	15.1	22.7
87	11.1	9.0	13.6



#44
 Trichloroethene
 Concen: 17.046 ug/l
 RT: 7.21 min Scan# 994
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
130	35389		
95	102.4	0.0	199.2









#45
 1,2-Dichloropropane
 Concen: 18.064 ug/l
 RT: 7.51 min Scan# 1043
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion: 63 Resp: 43926
 Ion Ratio Lower Upper
 63 100
 65 30.4 25.0 37.6



#46
 Dibromomethane
 Concen: 17.523 ug/l
 RT: 7.66 min Scan# 1067
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 93 Resp: 24743
 Ion Ratio Lower Upper
 93 100
 95 83.5 66.5 99.7
 174 103.1 82.1 123.1









#47

Bromodichloromethane
 Concen: 17.059 ug/l
 RT: 7.90 min Scan# 1106
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01




Tgt Ion	Resp	Lower	Upper
83	48050		
85	63.8	52.4	78.6
127	7.9	6.6	10.0




#48

Methyl methacrylate
 Concen: 18.720 ug/l
 RT: 7.77 min Scan# 1085
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
41	57434		
69	70.0	59.1	88.7
39	51.0	40.4	60.6









#49
 1,4-Dioxane
 Concen: 352.687 ug/l
 RT: 7.74 min Scan# 1080
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
88	21944		
43	38.0	28.6	42.8
58	80.5	61.7	92.5



#50
 Toluene-d8
 Concen: 44.571 ug/l
 RT: 8.71 min Scan# 1240
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
98	337759		
100	63.4	51.2	76.8









#51
 4-Methyl-2-Pentanone
 Concen: 94.162 ug/l
 RT: 8.64 min Scan# 1228
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
43	380605		
58	35.4	29.5	44.3



#52
 Toluene
 Concen: 17.463 ug/l
 RT: 8.78 min Scan# 1251
 Delta R.T. -0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
92	91160		
91	175.3	138.8	208.2









#53
 t-1,3-Dichloropropene
 Concen: 17.505 ug/l
 RT: 9.04 min Scan# 1294
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
75	55667		
77	30.6	25.0	37.6



#54
 cis-1,3-Dichloropropene
 Concen: 17.382 ug/l
 RT: 8.43 min Scan# 1194
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
75	61183		
77	32.9	24.6	36.8
39	57.6	40.6	60.8









#55
 1,1,2-Trichloroethane
 Concen: 17.673 ug/l
 RT: 9.21 min Scan# 1322
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01

Tgt Ion	Resp	Lower	Upper
97	100		
83	86.3	68.9	103.3
85	58.8	43.8	65.6
99	61.0	48.2	72.2




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 17.706 ug/l
 RT: 9.18 min Scan# 1316
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
69	100		
41	79.5	60.6	90.8
39	37.4	28.2	42.4



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0








#57
 1,3-Dichloropropane
 Concen: 17.533 ug/l
 RT: 9.37 min Scan# 1348
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10





Instrument :
 MSVOA_X
ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
76	65707		
78	32.5	25.5	38.3



#58
 2-Chloroethyl Vinyl ether
 Concen: 85.391 ug/l
 RT: 8.31 min Scan# 1174
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
63	172208		
106	24.1	19.7	29.5









#59
 2-Hexanone
 Concen: 92.551 ug/l
 RT: 9.49 min Scan# 1367
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion: 43 Resp: 293807
 Ion Ratio Lower Upper
 43 100
 58 50.0 25.9 77.7



#60
 Dibromochloromethane
 Concen: 16.929 ug/l
 RT: 9.58 min Scan# 1382
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 129 Resp: 35213
 Ion Ratio Lower Upper
 129 100
 127 78.9 38.4 115.1









#61
 1,2-Dibromoethane
 Concen: 17.484 ug/l
 RT: 9.67 min Scan# 1397
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
107	100		
109	95.7	75.2	112.8



#62
 4-Bromofluorobenzene
 Concen: 43.545 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
95	100		
174	81.3	0.0	161.6
176	78.1	0.0	159.2









#63
 Chlorobenzene-d5
 Concen: 50.000 ug/l
 RT: 10.11 min Scan# 1469
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
117	266547		
82	60.9	48.8	73.2
119	32.0	25.8	38.6



#64
 Tetrachloroethene
 Concen: 18.083 ug/l
 RT: 9.34 min Scan# 1342
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
164	33062		
166	129.7	101.2	151.8
129	93.4	74.3	111.5
131	91.9	71.0	106.6









#65
 Chlorobenzene
 Concen: 17.554 ug/l
 RT: 10.14 min Scan# 1474
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
112	93599		
114	31.4	25.5	38.3



#66
 1,1,1,2-Tetrachloroethane
 Concen: 17.377 ug/l
 RT: 10.22 min Scan# 1487
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
131	33591		
133	94.1	47.7	143.1
119	67.7	33.5	100.4









#67
 Ethyl Benzene
 Concen: 18.032 ug/l
 RT: 10.25 min Scan# 1492
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Instrument : MSVOA_X
 ClientSampled : VX0523MBS01

Tgt Ion: 91 Resp: 176949




Ion	Ratio	Lower	Upper
91	100		
106	29.7	24.0	36.0




#68
 m/p-Xylenes
 Concen: 36.058 ug/l
 RT: 10.36 min Scan# 1510
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 106 Resp: 129305

Ion	Ratio	Lower	Upper
106	100		
91	207.8	164.0	246.0









#69
 o-Xylene
 Concen: 17.552 ug/l
 RT: 10.69 min Scan# 1565
 Delta R.T. -0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion:106 Resp: 62101
 Ion Ratio Lower Upper
 106 100
 91 223.8 109.5 328.4



#70
 Styrene
 Concen: 17.526 ug/l
 RT: 10.71 min Scan# 1568
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion:104 Resp: 107775
 Ion Ratio Lower Upper
 104 100
 78 53.6 41.0 61.4
 103 55.9 43.8 65.6









#71
 Bromoform
 Concen: 16.979 ug/l
 RT: 10.85 min Scan# 1591
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
173	26673		
175	48.9	24.5	73.5
254	0.0	0.1	0.1#



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.08 min Scan# 1792
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
152	129886		
115	67.8	41.2	123.6
150	162.4	0.0	346.4









#73
 Isopropylbenzene
 Concen: 17.959 ug/l
 RT: 11.02 min Scan# 1618
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
105	169623		
120	25.5	12.8	38.3



#74
 N-amyl acetate
 Concen: 18.218 ug/l
 RT: 10.90 min Scan# 1598
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
43	100318		
70	36.4	30.0	45.0
55	22.2	17.9	26.9
61	22.0	18.4	27.6









#75
 1,1,2,2-Tetrachloroethane
 Concen: 17.617 ug/l
 RT: 11.27 min Scan# 1659
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
83	100		
131	9.3	4.6	13.8
85	63.6	32.0	96.2



#76
 1,2,3-Trichloropropane
 Concen: 23.205 ug/l
 RT: 11.29 min Scan# 1663
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
75	100		
77	33.2	22.1	66.1









#77
 Bromobenzene
 Concen: 17.585 ug/l
 RT: 11.26 min Scan# 1657
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument : MSVOA_X
 ClientSampled : VX0523MBS01


Tgt Ion	Resp	Lower	Upper
156	41226		
77	171.7	84.8	254.3
158	98.4	49.1	147.3



#78
 n-propylbenzene
 Concen: 17.803 ug/l
 RT: 11.36 min Scan# 1674
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
91	192041		
120	22.1	11.0	33.0









#79
 2-Chlorotoluene
 Concen: 17.321 ug/l
 RT: 11.42 min Scan# 1684
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion: 91 Resp: 115269
 Ion Ratio Lower Upper
 91 100
 126 32.4 16.3 48.9



#80
 1,3,5-Trimethylbenzene
 Concen: 17.933 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 105 Resp: 142619
 Ion Ratio Lower Upper
 105 100
 120 47.8 24.3 72.9









#81
 trans-1,4-Dichloro-2-butene
 Concen: 16.168 ug/l
 RT: 11.07 min Scan# 1627
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
75	19038		
75	100		
53	109.5	79.0	118.6
89	43.0	34.7	52.1



#82
 4-Chlorotoluene
 Concen: 17.758 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
91	133508		
91	100		
126	28.4	14.5	43.6









#83
 tert-Butylbenzene
 Concen: 17.160 ug/l
 RT: 11.77 min Scan# 1741
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
119	132756		
91	60.6	29.5	88.5
134	23.0	11.4	34.1



#84
 1,2,4-Trimethylbenzene
 Concen: 18.047 ug/l
 RT: 11.80 min Scan# 1747
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
105	144063		
120	43.4	22.0	66.0









#85
 sec-Butylbenzene
 Concen: 17.635 ug/l
 RT: 11.94 min Scan# 1770
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion: 105 Resp: 161747
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 17.894 ug/l
 RT: 12.06 min Scan# 1789
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion: 119 Resp: 147548
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.9 38.7
 91 24.1 11.8 35.4









#87
 1,3-Dichlorobenzene
 Concen: 17.540 ug/l
 RT: 12.02 min Scan# 1783
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
146	100		
111	40.2	20.0	59.9
148	63.8	31.8	95.3



#88
 1,4-Dichlorobenzene
 Concen: 17.554 ug/l
 RT: 12.10 min Scan# 1795
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
146	100		
111	40.5	19.6	58.8
148	63.7	31.8	95.3









#89
 n-Butylbenzene
 Concen: 17.689 ug/l
 RT: 12.39 min Scan# 1843
 Delta R.T. 0.01 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampleID :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
91	131831		
92	52.4	26.5	79.5
134	24.6	12.4	37.2



#90
 Hexachloroethane
 Concen: 16.482 ug/l
 RT: 12.60 min Scan# 1877
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
117	23154		
119	100		
201	87.9	42.3	126.9









#91
 1,2-Dichlorobenzene
 Concen: 17.251 ug/l
 RT: 12.39 min Scan# 1843
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
146	100		
111	41.4	20.8	62.4
148	64.6	31.9	95.9



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 18.731 ug/l
 RT: 13.00 min Scan# 1943
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
75	100		
155	78.8	42.0	126.0
157	98.6	53.2	159.6









#93
 1,2,4-Trichlorobenzene
 Concen: 17.639 ug/l
 RT: 13.65 min Scan# 2049
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10




Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01


Tgt Ion	Resp	Lower	Upper
180	100		
182	95.9	48.4	145.2
145	29.8	14.6	43.8



#94
 Hexachlorobutadiene
 Concen: 17.359 ug/l
 RT: 13.78 min Scan# 2071
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
225	100		
223	64.0	31.3	93.8
227	65.5	31.8	95.3









#95
 Naphthalene
 Concen: 17.638 ug/l
 RT: 13.83 min Scan# 2080
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Instrument :
 MSVOA_X
 ClientSampled :
 VX0523MBS01

Tgt Ion	Resp	Lower	Upper
128	167033		
127	13.5	10.4	15.6
129	11.0	8.6	13.0



#96
 1,2,3-Trichlorobenzene
 Concen: 17.664 ug/l
 RT: 14.02 min Scan# 2110
 Delta R.T. -0.00 min
 Lab File: VX009853.D
 Acq: 23 May 2019 13:10

Tgt Ion	Resp	Lower	Upper
180	52120		
182	95.8	47.8	143.4
145	31.4	15.4	46.4

