

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
 Data File : VX009851.D
 Acq On : 23 May 2019 12:24
 Operator : JC/SP
 Sample : K2974-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 BP-VPB202-EB-20190520

Quant Time: May 23 14:17:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

| Internal Standards | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|---------|-----------|
| 1) Pentafluorobenzene | 5.66 | 168 | 172871 | 50.00 | ug/l | 0.00 |
| 34) 1,4-Difluorobenzene | 6.86 | 114 | 275423 | 50.00 | ug/l | 0.00 |
| 63) Chlorobenzene-d5 | 10.11 | 117 | 242138 | 50.00 | ug/l | 0.00 |
| 72) 1,4-Dichlorobenzene-d4 | 12.08 | 152 | 101099 | 50.00 | ug/l | 0.00 |
| System Monitoring Compounds | | | | | | |
| 33) 1,2-Dichloroethane-d4 | 6.06 | 65 | 119171 | 50.50 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 101.00% | |
| 35) Dibromofluoromethane | 5.50 | 113 | 85384 | 49.31 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 98.62% | |
| 50) Toluene-d8 | 8.71 | 98 | 348451 | 49.85 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 99.70% | |
| 62) 4-Bromofluorobenzene | 11.13 | 95 | 113722 | 44.77 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 89.54% | |
| Target Compounds | | | | | | |
| 16) Acetone | 2.45 | 43 | 12220 | 9.644 | ug/l | Qvalue 92 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
 Data File : VX009851.D
 Acq On : 23 May 2019 12:24
 Operator : JC/SP
 Sample : K2974-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB202-EB-20190520

Quant Time: May 23 14:17:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

