

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
 Data File : VX009861.D
 Acq On : 23 May 2019 16:17
 Operator : JC/SP
 Sample : K2999-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-190521

Quant Time: May 24 04:35:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	175080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	279353	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244798	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	99671	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119268	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
35) Dibromofluoromethane	5.50	113	86022	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	8.71	98	357446	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
62) 4-Bromofluorobenzene	11.13	95	115799	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	20721	9.744	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052319\
Data File : VX009861.D
Acq On : 23 May 2019 16:17
Operator : JC/SP
Sample : K2999-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-190521

Quant Time: May 24 04:35:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

