

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052319\
 Data File : VX009864.D
 Acq On : 23 May 2019 17:28
 Operator : JC/SP
 Sample : K2976-03DL 250X
 Misc : 6.55g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SB-E3-4R(14-17)DL

Quant Time: May 31 08:14:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	183094	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	293742	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	265764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	120800	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	125651	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
35) Dibromofluoromethane	5.50	113	88826	48.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.20%	
50) Toluene-d8	8.71	98	372903	50.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.04%	
62) 4-Bromofluorobenzene	11.13	95	131755	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
Target Compounds						
73) Isopropylbenzene	11.02	105	13294	1.513	ug/l	97
78) n-propylbenzene	11.36	91	41298	4.117	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	58420	7.898	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	208839	28.129	ug/l	100
85) sec-Butylbenzene	11.94	105	24783	2.905	ug/l	99
86) p-Isopropyltoluene	12.06	119	21205	2.765	ug/l	98
89) n-Butylbenzene	12.38	91	33132	4.780	ug/l #	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052319\
Data File : VX009864.D
Acq On : 23 May 2019 17:28
Operator : JC/SP
Sample : K2976-03DL 250X
Misc : 6.55g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-SB-E3-4R(14-17)DL

Quant Time: May 31 08:14:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

