

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX052319\
 Data File : VX009867.D
 Acq On : 23 May 2019 18:38
 Operator : JC/SP
 Sample : K2976-20 10X
 Misc : 5.74g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP1-SB-SBR-3BR(17-20)

Quant Time: May 24 04:52:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	170779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	274635	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	111097	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116775	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
35) Dibromofluoromethane	5.50	113	83247	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
50) Toluene-d8	8.71	98	354301	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	11.13	95	144756	57.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.30%	
Target Compounds						
73) Isopropylbenzene	11.02	105	158082	19.567	ug/l	99
78) n-propylbenzene	11.36	91	278159	30.148	ug/l	98
83) tert-Butylbenzene	11.77	119	33626	5.081	ug/l	84
84) 1,2,4-Trimethylbenzene	11.80	105	153272	22.448	ug/l	99
85) sec-Butylbenzene	11.94	105	249969	31.862	ug/l	97
89) n-Butylbenzene	12.38	91	151087	23.701	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052319\
Data File : VX009867.D
Acq On : 23 May 2019 18:38
Operator : JC/SP
Sample : K2976-20 10X
Misc : 5.74g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP1-SB-SBR-3BR(17-20)

Quant Time: May 24 04:52:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

