

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028889.D
 Acq On : 23 May 2022 13:07
 Operator : JC/MD
 Sample : VX0523WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523WBL01

Quant Time: May 23 16:02:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	351411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	612618	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	583331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	270207	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	245418	50.948	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.900%
35) Dibromofluoromethane	5.391	113	206308	51.400	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.800%
50) Toluene-d8	8.653	98	729098	48.180	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.360%
62) 4-Bromofluorobenzene	11.079	95	287430	50.853	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
Data File : VX028889.D
Acq On : 23 May 2022 13:07
Operator : JC/MD
Sample : VX0523WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0523WBL01

Quant Time: May 23 16:02:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

