

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028894.D
 Acq On : 23 May 2022 15:12
 Operator : JC/MD
 Sample : N2970-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB01-20220518

Quant Time: May 24 05:33:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	362142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	639897	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	606368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	291228	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	244921	49.338	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.680%
35) Dibromofluoromethane	5.385	113	208031	49.620	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.240%
50) Toluene-d8	8.653	98	753098	47.645	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.280%
62) 4-Bromofluorobenzene	11.079	95	307832	52.141	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.280%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	40316	18.923	ug/l	93
20) Methylene Chloride	2.788	84	3494	0.782	ug/l	97
25) 2-Butanone	4.568	43	6143	1.837	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
Data File : VX028894.D
Acq On : 23 May 2022 15:12
Operator : JC/MD
Sample : N2970-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB01-20220518

Quant Time: May 24 05:33:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

