

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028895.D
 Acq On : 23 May 2022 15:35
 Operator : JC/MD
 Sample : N2970-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PDB-BLANK-20220519

Quant Time: May 24 05:33:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	354391	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	624736	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	592740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	282985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	242465	49.912	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.820%
35) Dibromofluoromethane	5.391	113	201337	49.188	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.380%
50) Toluene-d8	8.653	98	736733	47.740	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.480%
62) 4-Bromofluorobenzene	11.079	95	300585	52.149	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.300%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3632	1.742	ug/l	96
23) Vinyl Acetate	3.727	43	23745	2.191	ug/l	# 92
37) Ethyl Acetate	4.715	43	1464954	228.138	ug/l	99
48) Methyl methacrylate	7.696	41	6283	1.295	ug/l	91
52) Toluene	8.720	92	5041	0.444	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
Data File : VX028895.D
Acq On : 23 May 2022 15:35
Operator : JC/MD
Sample : N2970-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PDB-BLANK-20220519

Quant Time: May 24 05:33:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

