

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028900.D
 Acq On : 23 May 2022 17:32
 Operator : JC/MD
 Sample : N2958-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20220516

Quant Time: May 24 05:34:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	365805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	637962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	609626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	284800	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	239745	47.812	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.620%
35) Dibromofluoromethane	5.391	113	203218	48.619	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.240%
50) Toluene-d8	8.652	98	755062	47.914	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.820%
62) 4-Bromofluorobenzene	11.085	95	303148	51.503	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.000%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.946	59	28231	19.162	ug/l #	76
16) Acetone	2.379	43	4375	2.033	ug/l	94
20) Methylene Chloride	2.782	84	1369	0.303	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
Data File : VX028900.D
Acq On : 23 May 2022 17:32
Operator : JC/MD
Sample : N2958-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20220516

Quant Time: May 24 05:34:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

