

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028901.D
 Acq On : 23 May 2022 17:56
 Operator : JC/MD
 Sample : N2958-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20220517

Quant Time: May 24 05:35:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	354104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	623917	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	597513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	281155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	236640	48.752	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.500%
35) Dibromofluoromethane	5.391	113	201091	49.193	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.380%
50) Toluene-d8	8.653	98	738673	47.929	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.860%
62) 4-Bromofluorobenzene	11.085	95	300786	52.252	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.500%
Target Compounds						
16) Acetone	2.380	43	3488	1.674	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
Data File : VX028901.D
Acq On : 23 May 2022 17:56
Operator : JC/MD
Sample : N2958-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20220517

Quant Time: May 24 05:35:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

