

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028903.D
 Acq On : 23 May 2022 18:43
 Operator : JC/MD
 Sample : N2958-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01S-20220517

Quant Time: May 24 05:36:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	353174	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	615736	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	583724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	271211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	235095	48.561	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.120%
35) Dibromofluoromethane	5.391	113	201214	49.877	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.760%
50) Toluene-d8	8.653	98	729568	47.967	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.940%
62) 4-Bromofluorobenzene	11.085	95	290728	51.176	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.360%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3978	1.915	ug/l	94
29) Tetrahydrofuran	5.019	42	11508	5.429	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
Data File : VX028903.D
Acq On : 23 May 2022 18:43
Operator : JC/MD
Sample : N2958-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01S-20220517

Quant Time: May 24 05:36:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

