

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028905.D
 Acq On : 23 May 2022 19:29
 Operator : JC/MD
 Sample : N2983-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 16S-20220520

Quant Time: May 24 05:36:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	338738	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	599968	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	579400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	276562	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	234500	50.503	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.000%
35) Dibromofluoromethane	5.391	113	196943	50.101	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.200%
50) Toluene-d8	8.653	98	718760	48.499	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.000%
62) 4-Bromofluorobenzene	11.085	95	289776	52.349	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.700%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	2898	1.454	ug/l	# 76
32) 1,1,1-Trichloroethane	5.397	97	1894	0.287	ug/l	# 48
44) Trichloroethene	7.129	130	2923	0.635	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052322\
 Data File : VX028905.D
 Acq On : 23 May 2022 19:29
 Operator : JC/MD
 Sample : N2983-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 16S-20220520

Quant Time: May 24 05:36:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

