

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052323\
 Data File : VX035812.D
 Acq On : 23 May 2023 13:58
 Operator : JC/MD
 Sample : VX0523WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0523WBL01

Quant Time: May 24 05:48:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	32153	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	176822	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	155676	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	80554	30.152	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.500%
60) 4-Bromofluorobenzene	11.079	95	71847	27.504	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.667%
63) Toluene-d8	8.647	98	209461	30.435	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.467%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052323\
Data File : VX035812.D
Acq On : 23 May 2023 13:58
Operator : JC/MD
Sample : VX0523WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0523WBL01

Quant Time: May 24 05:48:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

