

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052323\
 Data File : VX035816.D
 Acq On : 23 May 2023 16:06
 Operator : JC/MD
 Sample : 02872-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL-DSN-001

Quant Time: May 24 05:49:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	32405	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	178146	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	156114	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	80972	30.073	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.233%
60) 4-Bromofluorobenzene	11.079	95	74847	28.572	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.233%
63) Toluene-d8	8.647	98	208995	30.282	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.933%
Target Compounds						
					Qvalue	
15) Acetone	2.398	58	35286	102.526	ug/l	92
25) Chloroform	5.087	83	9014	2.369	ug/l	99
41) Bromodichloromethane	7.824	83	2988	1.160	ug/l #	95
53) Dibromochloromethane	9.525	129	1401	0.749	ug/l	98
56) Bromoform	10.799	173	517	0.427	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052323\
Data File : VX035816.D
Acq On : 23 May 2023 16:06
Operator : JC/MD
Sample : 02872-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-DSN-001

Quant Time: May 24 05:49:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration

