

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052323\
 Data File : VX035817.D
 Acq On : 23 May 2023 16:29
 Operator : JC/MD
 Sample : 02872-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL-DSN-002

Quant Time: May 24 05:50:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 23 11:36:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	4.904	128	31881	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.763	114	188523	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	165924	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	80795	30.500	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.667%
60) 4-Bromofluorobenzene	11.079	95	77963	28.002	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.333%
63) Toluene-d8	8.647	98	222100	30.279	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.933%
Target Compounds						
					Qvalue	
15) Acetone	2.398	58	10645	31.438	ug/l	81
25) Chloroform	5.099	83	18318	4.893	ug/l	93
41) Bromodichloromethane	7.824	83	3237	1.188	ug/l #	91
53) Dibromochloromethane	9.525	129	1229	0.621	ug/l	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052323\
Data File : VX035817.D
Acq On : 23 May 2023 16:29
Operator : JC/MD
Sample : 02872-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-DSN-002

Quant Time: May 24 05:50:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X052323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 23 11:36:37 2023
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 30.000 ug/l

RT: 4.904 min Scan# 61

Delta R.T. 0.013 min

Lab File: VX035817.D

Acq: 23 May 2023 16:29

Instrument :

MSVOA_X

ClientSampleId :

OUTFALL-DSN-002





Tgt Ion:128 Resp: 31881

Ion Ratio Lower Upper

128 100

49 175.8 0.0 428.8

130 130.9 0.0 310.8



#15

Acetone

Concen: 31.438 ug/l

RT: 2.398 min Scan# 215

Delta R.T. -0.006 min

Lab File: VX035817.D




Acq: 23 May 2023 16:29


Tgt Ion: 58 Resp: 10645

Ion Ratio Lower Upper

58 100

43 298.1 270.8 406.2









#25
Chloroform
Concen: 4.893 ug/l
RT: 5.099 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX035817.D
Acq: 23 May 2023 16:29




Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-DSN-002


Tgt Ion: 83 Resp: 18318
Ion Ratio Lower Upper
83 100
85 69.8 51.2 76.8



#27
1,2-Dichloroethane-d4
Concen: 30.500 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX035817.D
Acq: 23 May 2023 16:29

Tgt Ion: 65 Resp: 80795
Ion Ratio Lower Upper
65 100
67 53.4 40.9 61.3





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. 0.006 min

Lab File: VX035817.D

Acq: 23 May 2023 16:29

Instrument :

MSVOA_X

ClientSampleId :

OUTFALL-DSN-002





Tgt Ion:114 Resp: 188523

Ion Ratio Lower Upper

114 100

63 21.8 18.0 27.0

88 16.0 13.3 19.9



#41

Bromodichloromethane

Concen: 1.188 ug/l

RT: 7.824 min Scan# 1105

Delta R.T. 0.006 min

Lab File: VX035817.D

Acq: 23 May 2023 16:29




Tgt Ion: 83 Resp: 3237


Ion Ratio Lower Upper

83 100

85 70.7 51.1 76.7

127 14.1 8.2 12.4#









#53
Dibromochloromethane
Concen: 0.621 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.006 min
Lab File: VX035817.D
Acq: 23 May 2023 16:29




Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-DSN-002


Tgt Ion:129 Resp: 1229
Ion Ratio Lower Upper
129 100
127 66.5 39.1 117.5



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX035817.D
Acq: 23 May 2023 16:29

Tgt Ion:117 Resp: 165924
Ion Ratio Lower Upper
117 100
82 57.2 45.8 68.8
119 31.8 25.4 38.0









#60
4-Bromofluorobenzene
Concen: 28.002 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX035817.D
Acq: 23 May 2023 16:29

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-DSN-002

Tgt Ion: 95 Resp: 77963
Ion Ratio Lower Upper
95 100
174 74.0 0.2 0.4#
176 70.1 0.2 0.2#



#63
Toluene-d8
Concen: 30.279 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX035817.D
Acq: 23 May 2023 16:29

Tgt Ion: 98 Resp: 222100
Ion Ratio Lower Upper
98 100
100 67.2 52.1 78.1

