

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052324\
 Data File : VX041622.D
 Acq On : 23 May 2024 12:39
 Operator : JC/MD
 Sample : PB161061TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: May 24 05:50:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	141737	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	212704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105727	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79120	51.277	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.560%	
35) Dibromofluoromethane	5.385	113	66491	46.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.260%	
50) Toluene-d8	8.647	98	254394	55.761	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 111.520%	
62) 4-Bromofluorobenzene	11.079	95	96365	54.564	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.120%	


Target Compounds				Qvalue		
3) Chloromethane	1.301	50	250	0.207	ug/l #	43
5) Bromomethane	1.599	94	728	1.011	ug/l #	74
6) Chloroethane	1.673	64	423	0.793	ug/l #	43
11) Tert butyl alcohol	2.947	59	5276	17.083	ug/l #	78
13) Acrolein	2.239	56	221	0.902	ug/l	95
16) Acetone	2.392	43	28546	38.705	ug/l	96
17) Carbon Disulfide	2.508	76	684	0.265	ug/l #	76
18) Methyl Acetate	2.715	43	2084	1.320	ug/l	93
20) Methylene Chloride	2.782	84	10414	7.774	ug/l	90
25) 2-Butanone	4.617	43	1771	1.773	ug/l	92
29) Tetrahydrofuran	5.056	42	620	0.949	ug/l #	45
31) Cyclohexane	5.550	56	2629	1.412	ug/l #	1
36) 1,1-Dichloropropene	5.550	75	9328	5.412	ug/l #	50
37) Ethyl Acetate	4.745	43	2400	1.116	ug/l #	90
38) Carbon Tetrachloride	5.550	117	12430	6.295	ug/l #	14
41) Methacrylonitrile	5.044	41	418	0.378	ug/l #	54
43) Isopropyl Acetate	6.348	43	119253	36.880	ug/l	99
48) Methyl methacrylate	7.799	41	109844	74.104	ug/l #	48
49) 1,4-Dioxane	7.690	88	27	0.842	ug/l #	1
51) 4-Methyl-2-Pentanone	8.549	43	93035	47.949	ug/l #	50
53) t-1,3-Dichloropropene	8.952	75	957	2.906	ug/l #	43
54) cis-1,3-Dichloropropene	8.549	75	42002	22.686	ug/l #	60
56) Ethyl methacrylate	9.250	69	892	0.460	ug/l #	1
57) 1,3-Dichloropropane	9.482	76	5484	2.625	ug/l #	43
59) 2-Hexanone	9.482	43	81098	52.270	ug/l #	52
68) m/p-Xylenes	10.305	106	655	0.229	ug/l	89
74) N-amyl acetate	10.817	43	970	0.344	ug/l #	49
76) 1,2,3-Trichloropropane	11.079	75	49674	28.830	ug/l #	38
81) trans-1,4-Dichloro-2-b...	11.079	75	49674	78.528	ug/l #	5
93) 1,2,4-Trichlorobenzene	13.591	180	683	0.279	ug/l	91
95) Naphthalene	13.780	128	3017	0.416	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	653	0.261	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052324\
Data File : VX041622.D
Acq On : 23 May 2024 12:39
Operator : JC/MD
Sample : PB161061TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: May 24 05:50:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 141737
Ion Ratio Lower Upper
168 100
99 49.1 56.6 85.0#



#3
Chloromethane
Concen: 0.207 ug/l
RT: 1.301 min Scan# 35
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 50 Resp: 250
Ion Ratio Lower Upper
50 100
52 0.0 25.5 38.3#









#5
Bromomethane
Concen: 1.011 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 728
Ion Ratio Lower Upper
94 100
96 68.0 74.6 111.8#



#6
Chloroethane
Concen: 0.793 ug/l
RT: 1.673 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 64 Resp: 423
Ion Ratio Lower Upper
64 100
66 0.0 25.1 37.7#









#11
Tert butyl alcohol
Concen: 17.083 ug/l
RT: 2.947 min Scan# 305
Delta R.T. -0.073 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 59 Resp: 5276
Ion Ratio Lower Upper
59 100
57 0.0 6.0 9.0#



#13
Acrolein
Concen: 0.902 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39


Tgt Ion: 56 Resp: 221
Ion Ratio Lower Upper
56 100
55 75.6 57.1 85.7








#16
Acetone
Concen: 38.705 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 28546
Ion Ratio Lower Upper
43 100
58 29.4 21.9 32.9



#17
Carbon Disulfide
Concen: 0.265 ug/l
RT: 2.508 min Scan# 233
Delta R.T. 0.007 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 76 Resp: 684
Ion Ratio Lower Upper
76 100
78 0.0 7.0 10.6#









#18
Methyl Acetate
Concen: 1.320 ug/l
RT: 2.715 min Scan# 2084
Delta R.T. 0.006 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 2084
Ion Ratio Lower Upper
43 100
74 18.9 17.8 26.8



#20
Methylene Chloride
Concen: 7.774 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.006 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 84 Resp: 10414
Ion Ratio Lower Upper
84 100
49 114.2 103.6 155.4
51 35.5 33.8 50.6
86 59.6 49.4 74.0





#25

2-Butanone

Concen: 1.773 ug/l

RT: 4.617 min Scan# 51

Delta R.T. 0.049 min

Lab File: VX041622.D

Acq: 23 May 2024 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 1771

Ion Ratio Lower Upper

43 100

72 28.1 19.4 29.2



#29

Tetrahydrofuran

Concen: 0.949 ug/l

RT: 5.056 min Scan# 651

Delta R.T. 0.043 min

Lab File: VX041622.D

Acq: 23 May 2024 12:39




Tgt Ion: 42 Resp: 620


Ion Ratio Lower Upper

42 100

72 0.0 34.2 51.4#


71 14.7 32.2 48.2#








#31
Cyclohexane
Concen: 1.412 ug/l
RT: 5.550 min Scan# 718
Delta R.T. 0.086 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 2629
Ion Ratio Lower Upper
56 100
69 163.6 23.4 35.2#
84 157.0 65.9 98.9#



#33
1,2-Dichloroethane-d4
Concen: 51.277 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 65 Resp: 79120
Ion Ratio Lower Upper
65 100
67 50.1 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:114 Resp: 212704
Ion Ratio Lower Upper
114 100
63 18.7 0.0 45.0
88 14.4 0.0 36.6



#35
Dibromofluoromethane
Concen: 46.128 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion:113 Resp: 66491
Ion Ratio Lower Upper
113 100
111 102.2 82.9 124.3
192 25.1 13.3 19.9#









#36
1,1-Dichloropropene
Concen: 5.412 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.140 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 9328
Ion Ratio Lower Upper
75 100
110 8.8 17.0 50.9#
77 0.0 24.7 37.1#



#37
Ethyl Acetate
Concen: 1.116 ug/l
RT: 4.745 min Scan# 600
Delta R.T. 0.025 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39


Tgt Ion: 43 Resp: 2400
Ion Ratio Lower Upper
43 100
61 10.3 10.6 16.0#
70 5.8 8.3 12.5#








#38
Carbon Tetrachloride
Concen: 6.295 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.122 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 12430
Ion Ratio Lower Upper
117 100
119 3.7 78.2 117.2#
121 0.0 24.4 36.6#



#41
Methacrylonitrile
Concen: 0.378 ug/l
RT: 5.044 min Scan# 649
Delta R.T. 0.116 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 41 Resp: 418
Ion Ratio Lower Upper
41 100
39 56.9 45.2 67.8
67 0.0 52.6 78.8#
52 0.0 22.7 34.1#









#43
Isopropyl Acetate
Concen: 36.880 ug/l
RT: 6.348 min Scan# 80
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 119253
Ion Ratio Lower Upper
43 100
61 18.8 15.2 22.8
87 13.3 9.8 14.6



#48
Methyl methacrylate
Concen: 74.104 ug/l
RT: 7.799 min Scan# 1101
Delta R.T. 0.104 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39


Tgt Ion: 41 Resp: 109844
Ion Ratio Lower Upper
41 100
69 0.1 59.2 88.8#
39 53.9 47.7 71.5








#49
1,4-Dioxane
Concen: 0.842 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.018 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 88 Resp: 27
Ion Ratio Lower Upper
88 100
43 0.0 0.0 0.0
58 2659.3 60.1 90.1#



#50
Toluene-d8
Concen: 55.761 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 98 Resp: 254394
Ion Ratio Lower Upper
98 100
100 66.2 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 47.949 ug/l
RT: 8.549 min Scan# 11
Delta R.T. -0.030 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 93035
Ion Ratio Lower Upper
43 100
58 7.8 30.4 45.6#



#53
t-1,3-Dichloropropene
Concen: 2.906 ug/l
RT: 8.952 min Scan# 1290
Delta R.T. -0.024 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 75 Resp: 957
Ion Ratio Lower Upper
75 100
77 0.0 25.7 38.5#









#54
 cis-1,3-Dichloropropene
 Concen: 22.686 ug/l
 RT: 8.549 min Scan# 1194
 Delta R.T. 0.183 min
 Lab File: VX041622.D
 Acq: 23 May 2024 12:39




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 42002
 Ion Ratio Lower Upper
 75 100
 77 0.6 25.3 37.9#
 39 34.6 46.7 70.1#



#56
 Ethyl methacrylate
 Concen: 0.460 ug/l
 RT: 9.250 min Scan# 1339
 Delta R.T. 0.134 min
 Lab File: VX041622.D
 Acq: 23 May 2024 12:39

Tgt Ion: 69 Resp: 892
 Ion Ratio Lower Upper
 69 100
 41 7039.0 56.8 85.2#
 39 2921.7 32.0 48.0#





#57

1,3-Dichloropropane

Concen: 2.625 ug/l

RT: 9.482 min Scan# 1377

Delta R.T. 0.177 min

Lab File: VX041622.D

Acq: 23 May 2024 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 5484

Ion Ratio Lower Upper

76 100

78 0.4 26.1 39.1#



#59

2-Hexanone

Concen: 52.270 ug/l

RT: 9.482 min Scan# 1377

Delta R.T. 0.049 min

Lab File: VX041622.D




Acq: 23 May 2024 12:39


Tgt Ion: 43 Resp: 81098

Ion Ratio Lower Upper

43 100

58 18.7 26.2 78.5#









#62
4-Bromofluorobenzene
Concen: 54.564 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 95 Resp: 96365
Ion Ratio Lower Upper
95 100
174 97.3 0.0 135.6
176 95.3 0.0 131.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39

Tgt Ion: 117 Resp: 213430
Ion Ratio Lower Upper
117 100
82 52.4 46.2 69.4
119 32.8 25.5 38.3





#68

m/p-Xylenes

Concen: 0.229 ug/l

RT: 10.305 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX041622.D

Acq: 23 May 2024 12:39

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 655

Ion Ratio Lower Upper

106 100

91 190.1 165.5 248.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX041622.D

Acq: 23 May 2024 12:39




Tgt Ion:152 Resp: 105727


Ion Ratio Lower Upper

152 100

115 54.5 43.3 129.9

150 157.2 0.0 345.0





#74

N-amy1 acetate

Concen: 0.344 ug/l

RT: 10.817 min Scan# 1

Delta R.T. -0.024 min


Lab File: VX041622.D

Acq: 23 May 2024 12:39

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 43 Resp: 970




Ion	Ratio	Lower	Upper
-----	-------	-------	-------

43	100		
----	-----	--	--

70	0.0	32.9	49.3#
----	-----	------	-------

55	9.7	19.3	28.9#
----	-----	------	-------

61	0.0	18.8	28.2#
----	-----	------	-------



#76

1,2,3-Trichloropropane

Concen: 28.830 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.158 min

Lab File: VX041622.D




Acq: 23 May 2024 12:39


Tgt Ion: 75 Resp: 49674

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

75	100		
----	-----	--	--


77	0.9	19.1	57.5#
----	-----	------	-------








#81
trans-1,4-Dichloro-2-butene
Concen: 78.528 ug/l
RT: 11.079 min Scan# 1
Delta R.T. 0.061 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 75 Resp: 49674
Ion Ratio Lower Upper
75 100
53 0.0 90.5 135.7#
89 0.0 36.9 55.3#




#93
1,2,4-Trichlorobenzene
Concen: 0.279 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.006 min
Lab File: VX041622.D
Acq: 23 May 2024 12:39



Tgt Ion: 180 Resp: 683
Ion Ratio Lower Upper
180 100
182 89.2 48.3 144.8
145 28.6 18.0 54.0





#95

Naphthalene

Concen: 0.416 ug/l

RT: 13.780 min Scan# 2081

Delta R.T. 0.006 min

Lab File: VX041622.D





Acq: 23 May 2024 12:39

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:128	Resp:	3017
Ion Ratio	Lower	Upper
128	100	
127	12.7	10.3 15.5
129	11.0	8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 0.261 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX041622.D

Acq: 23 May 2024 12:39

Tgt Ion:180	Resp:	653
Ion Ratio	Lower	Upper
180	100	
182	98.2	46.9 140.6
145	30.8	17.8 53.5

