

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052324\
 Data File : VX041623.D
 Acq On : 23 May 2024 13:02
 Operator : JC/MD
 Sample : P2494-37
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 17

Quant Time: May 24 05:50:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147827	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	223937	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102748	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87157	54.158	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.320%	
35) Dibromofluoromethane	5.385	113	71217	46.929	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.860%	
50) Toluene-d8	8.647	98	263827	54.927	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.860%	
62) 4-Bromofluorobenzene	11.079	95	97150	52.250	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.500%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	31068	40.563	ug/l	93
18) Methyl Acetate	2.709	43	3204	1.946	ug/l	98
20) Methylene Chloride	2.788	84	7445	5.329	ug/l	92
37) Ethyl Acetate	4.745	43	3067	1.355	ug/l #	79
43) Isopropyl Acetate	6.342	43	115985	34.070	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052324\
Data File : VX041623.D
Acq On : 23 May 2024 13:02
Operator : JC/MD
Sample : P2494-37
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
17

Quant Time: May 24 05:50:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

