

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052324\  
 Data File : VX041633.D  
 Acq On : 23 May 2024 16:55  
 Operator : JC/MD  
 Sample : P2600-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 240522007-02-TRIP-BLANK

Quant Time: May 24 06:45:32 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X051524W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu May 16 05:13:13 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Bromochloromethane       | 4.904  | 128   | 26323    | 30.000   | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene     | 6.763  | 114   | 143371   | 30.000   | ug/l   | 0.00     |
| 57) Chlorobenzene-d5        | 10.055 | 117   | 126798   | 30.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 27) 1,2-Dichloroethane-d4   | 5.958  | 65    | 61594    | 29.714   | ug/l   | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 110 | Recovery | =      | 99.033%  |
| 60) 4-Bromofluorobenzene    | 11.079 | 95    | 55783    | 28.182   | ug/l   | 0.00     |
| Spiked Amount               | 30.000 | Range | 63 - 112 | Recovery | =      | 93.933%  |
| 63) Toluene-d8              | 8.647  | 98    | 162747   | 29.943   | ug/l   | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 112 | Recovery | =      | 99.800%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 15) Acetone                 | 2.392  | 58    | 988      | 4.135    | ug/l   | 72       |
| 25) Chloroform              | 5.105  | 83    | 4841     | 1.612    | ug/l   | 95       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052324\  
Data File : VX041633.D  
Acq On : 23 May 2024 16:55  
Operator : JC/MD  
Sample : P2600-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
240522007-02-TRIP-BLANK

Quant Time: May 24 06:45:32 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X051524W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Thu May 16 05:13:13 2024  
Response via : Initial Calibration

