



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052324\
Data File : VX041613.D
Acq On : 23 May 2024 08:23
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Title : SW846 8260
Last Update : Wed May 08 06:03:16 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	8110	PASS
75	95	30	60	51.6	21152	PASS
95	95	100	100	100.0	40971	PASS
96	95	5	9	6.7	2726	PASS
173	174	0.00	2	0.5	180	PASS
174	95	50	100	96.8	39667	PASS
175	174	5	9	7.3	2907	PASS
176	174	95	101	96.7	38373	PASS
177	176	5	9	6.3	2409	PASS