

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052419\
 Data File : VX009884.D
 Acq On : 24 May 2019 14:14
 Operator : JC/SP
 Sample : K3036-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FEQC-TB-052319

Quant Time: May 25 01:12:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178771	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	285409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	257986	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112618	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122602	50.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.48%	
35) Dibromofluoromethane	5.50	113	86952	48.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.92%	
50) Toluene-d8	8.71	98	364297	50.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.58%	
62) 4-Bromofluorobenzene	11.13	95	123303	46.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052419\
Data File : VX009884.D
Acq On : 24 May 2019 14:14
Operator : JC/SP
Sample : K3036-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FEQC-TB-052319

Quant Time: May 25 01:12:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

