

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052419\
 Data File : VX009889.D
 Acq On : 24 May 2019 16:10
 Operator : JC/SP
 Sample : K3032-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0657

Quant Time: May 25 01:26:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	177509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	286132	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252853	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110127	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122117	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
35) Dibromofluoromethane	5.50	113	88032	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
50) Toluene-d8	8.71	98	362374	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
62) 4-Bromofluorobenzene	11.13	95	123573	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
Target Compounds						
4) Vinyl Chloride	1.41	62	29456	13.526	ug/l	98
12) 1,1-Dichloroethene	2.37	96	3831	2.300	ug/l	100
16) Acetone	2.44	43	43076	33.106	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	248963	137.831	ug/l	96
25) 2-Butanone	4.68	43	8515	4.211	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	760290	352.637	ug/l	91
44) Trichloroethene	7.22	130	1970136	990.288	ug/l	98
65) Chlorobenzene	10.14	112	3683	0.728	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052419\
Data File : VX009889.D
Acq On : 24 May 2019 16:10
Operator : JC/SP
Sample : K3032-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0657

Quant Time: May 25 01:26:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

