

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052419\
 Data File : VX009892.D
 Acq On : 24 May 2019 17:21
 Operator : JC/SP
 Sample : K3032-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0654

Quant Time: May 25 01:33:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	176337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	281122	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251808	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110767	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122179	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
35) Dibromofluoromethane	5.50	113	86724	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	8.71	98	362621	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	11.13	95	121026	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
Target Compounds						
4) Vinyl Chloride	1.41	62	38717	17.897	ug/l	96
16) Acetone	2.44	43	5402	4.179	ug/l	94
27) cis-1,2-Dichloroethene	4.59	96	9465	4.419	ug/l	89
44) Trichloroethene	7.23	130	1047	0.536	ug/l	63
52) Toluene	8.79	92	226660	46.118	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052419\
Data File : VX009892.D
Acq On : 24 May 2019 17:21
Operator : JC/SP
Sample : K3032-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
S32-0654

Quant Time: May 25 01:33:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

