

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX052419\
 Data File : VX009893.D
 Acq On : 24 May 2019 17:44
 Operator : JC/SP
 Sample : K3032-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-ER-2019-1

Quant Time: May 25 01:35:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	174128	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	279823	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	250219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108223	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119362	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
35) Dibromofluoromethane	5.50	113	86405	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
50) Toluene-d8	8.71	98	355343	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.13	95	119720	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	2170	1.016	ug/l #	89
16) Acetone	2.45	43	4320	3.385	ug/l	100
20) Methylene Chloride	2.85	84	13608	6.585	ug/l	95
27) cis-1,2-Dichloroethene	4.61	96	725	0.343	ug/l	82
29) Tetrahydrofuran	5.13	42	79679	62.427	ug/l	96
44) Trichloroethene	7.21	130	1163	0.598	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX052419\
Data File : VX009893.D
Acq On : 24 May 2019 17:44
Operator : JC/SP
Sample : K3032-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-ER-2019-1

Quant Time: May 25 01:35:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

