

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052421\
 Data File : VX022161.D
 Acq On : 24 May 2021 20:32
 Operator : JC/MD
 Sample : M2472-12MS
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-938-940MS

Quant Time: May 25 02:10:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75340	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	153323	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	140860	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68489	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61773	53.15	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.30%	
35) Dibromofluoromethane	5.397	113	49368	50.77	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.54%	
50) Toluene-d8	8.653	98	187951	50.18	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.36%	
62) 4-Bromofluorobenzene	11.085	95	74891	54.10	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.20%	

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	44236	48.30	ug/l	99
3) Chloromethane	1.295	50	49762	50.37	ug/l	97
4) Vinyl Chloride	1.374	62	53720	51.39	ug/l	100
5) Bromomethane	1.593	94	13064	21.43	ug/l	97
6) Chloroethane	1.660	64	14310	22.99	ug/l	100
7) Trichlorofluoromethane	1.837	101	79631	52.98	ug/l	96
8) Diethyl Ether	2.142	74	34219	52.93	ug/l	88
9) 1,1,2-Trichlorotrifluo...	2.300	101	43977	49.27	ug/l	90
10) Methyl Iodide	2.428	142	30545	31.84	ug/l	95
11) Tert butyl alcohol	2.947	59	4988	18.77	ug/l	# 73
12) 1,1-Dichloroethene	2.294	96	42871	49.65	ug/l	87
13) Acrolein	2.252	56	43362	245.96	ug/l	99
14) Allyl chloride	2.648	41	80646	50.85	ug/l	96
15) Acrylonitrile	3.093	53	158781	303.42	ug/l	100
16) Acetone	2.416	43	140759	278.16	ug/l	98
17) Carbon Disulfide	2.483	76	105991	46.87	ug/l	99
18) Methyl Acetate	2.727	43	82736	61.48	ug/l	93
19) Methyl tert-butyl Ether	3.136	73	166361	56.24	ug/l	99
20) Methylene Chloride	2.782	84	57584	51.25	ug/l	94
21) trans-1,2-Dichloroethene	3.081	96	50858	54.15	ug/l	99
22) Diisopropyl ether	3.788	45	171218	56.05	ug/l	# 84
23) Vinyl Acetate	3.733	43	824906	280.48	ug/l	97
24) 1,1-Dichloroethane	3.605	63	100046	55.81	ug/l	98
25) 2-Butanone	4.599	43	241218	294.64	ug/l	94
26) 2,2-Dichloropropane	4.471	77	68112	42.80	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	62656	54.86	ug/l	98
28) Bromochloromethane	4.910	49	43253	49.74	ug/l	86
29) Tetrahydrofuran	5.038	42	141770	290.95	ug/l	91
30) Chloroform	5.105	83	100508	54.27	ug/l	96
31) Cyclohexane	5.465	56	74734	51.17	ug/l	97
32) 1,1,1-Trichloroethane	5.385	97	85052	53.74	ug/l	99
36) 1,1-Dichloropropene	5.690	75	72121	50.87	ug/l	98
37) Ethyl Acetate	4.745	43	93968	54.07	ug/l	96
38) Carbon Tetrachloride	5.672	117	70023	50.49	ug/l	97
39) Methylcyclohexane	7.379	83	75946	47.23	ug/l	98
40) Benzene	6.038	78	226176	51.85	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052421\
 Data File : VX022161.D
 Acq On : 24 May 2021 20:32
 Operator : JC/MD
 Sample : M2472-12MS
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-938-940MS

Quant Time: May 25 02:10:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.946	41	50472	54.28	ug/l	95
42) 1,2-Dichloroethane	6.099	62	79940	52.31	ug/l	98
43) Isopropyl Acetate	6.361	43	156575	55.08	ug/l	94
44) Trichloroethene	7.129	130	53791	49.62	ug/l	100
45) 1,2-Dichloropropane	7.434	63	60610	54.47	ug/l	98
46) Dibromomethane	7.586	93	39864	52.80	ug/l #	87
47) Bromodichloromethane	7.830	83	78542	51.99	ug/l	99
48) Methyl methacrylate	7.702	41	70945	54.15	ug/l	93
49) 1,4-Dioxane	7.690	88	27785	1089.29	ug/l	98
51) 4-Methyl-2-Pentanone	8.592	43	497485	292.15	ug/l	93
52) Toluene	8.720	92	143788	52.05	ug/l	96
53) t-1,3-Dichloropropene	8.982	75	87418	50.13	ug/l	97
54) cis-1,3-Dichloropropene	8.373	75	95344	50.86	ug/l	95
55) 1,1,2-Trichloroethane	9.159	97	60112	53.60	ug/l	99
56) Ethyl methacrylate	9.128	69	99427	54.86	ug/l #	90
57) 1,3-Dichloropropane	9.317	76	103342	53.97	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.251	63	253183	258.27	ug/l	97
59) 2-Hexanone	9.439	43	386199	301.55	ug/l	92
60) Dibromochloromethane	9.525	129	57920	53.09	ug/l	98
61) 1,2-Dibromoethane	9.616	107	60917	54.56	ug/l	99
64) Tetrachloroethene	9.275	164	44426	48.04	ug/l	92
65) Chlorobenzene	10.086	112	149763	51.07	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.171	131	53375	52.38	ug/l	97
67) Ethyl Benzene	10.195	91	276828	51.32	ug/l	100
68) m/p-Xylenes	10.305	106	211315	103.65	ug/l	100
69) o-Xylene	10.647	106	102983	51.94	ug/l	99
70) Styrene	10.659	104	176203	54.25	ug/l	98
71) Bromoform	10.805	173	39218	50.22	ug/l #	98
73) Isopropylbenzene	10.964	105	270291	46.63	ug/l	100
74) N-amyl acetate	10.848	43	149271	54.25	ug/l	92
75) 1,1,2,2-Tetrachloroethane	11.220	83	105660	54.99	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	117189	69.36	ug/l	91
77) Bromobenzene	11.201	156	60173	48.96	ug/l	90
78) n-propylbenzene	11.311	91	325095	47.88	ug/l	98
79) 2-Chlorotoluene	11.366	91	193298	46.94	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	236827	49.29	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.024	75	34114	47.84	ug/l	95
82) 4-Chlorotoluene	11.457	91	226475	47.73	ug/l	97
83) tert-Butylbenzene	11.719	119	227264	49.32	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	242826	50.21	ug/l	99
85) sec-Butylbenzene	11.896	105	286791	48.98	ug/l	99
86) p-Isopropyltoluene	12.012	119	243653	49.27	ug/l	97
87) 1,3-Dichlorobenzene	11.976	146	117753	49.69	ug/l	96
88) 1,4-Dichlorobenzene	12.043	146	121636	49.66	ug/l	96
89) n-Butylbenzene	12.335	91	231043	50.51	ug/l	98
90) Hexachloroethane	12.543	117	39962	48.89	ug/l	76
91) 1,2-Dichlorobenzene	12.341	146	118028	51.85	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	23372	56.85	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	72092	52.49	ug/l	97
94) Hexachlorobutadiene	13.731	225	27070	50.49	ug/l	99
95) Naphthalene	13.780	128	301381	56.26	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	73076	52.78	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052421\
Data File : VX022161.D
Acq On : 24 May 2021 20:32
Operator : JC/MD
Sample : M2472-12MS
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS

Quant Time: May 25 02:10:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:168 Resp: 75340
Ion Ratio Lower Upper
168 100
99 71.0 45.8 68.6#



#2
Dichlorodifluoromethane
Concen: 48.30 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.001 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 85 Resp: 44236
Ion Ratio Lower Upper
85 100
87 32.1 16.4 49.2









#3
Chloromethane
Concen: 50.37 ug/l
RT: 1.295 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 50 Resp: 49762
Ion Ratio Lower Upper
50 100
52 31.1 26.4 39.6



#4
Vinyl Chloride
Concen: 51.39 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 62 Resp: 53720
Ion Ratio Lower Upper
62 100
64 32.2 25.7 38.5









#5
Bromomethane
Concen: 21.43 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.019 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 94 Resp: 13064
Ion Ratio Lower Upper
94 100
96 98.3 76.4 114.6



#6
Chloroethane
Concen: 22.99 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.025 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 64 Resp: 14310
Ion Ratio Lower Upper
64 100
66 31.8 25.4 38.0









#7
 Trichlorofluoromethane
 Concen: 52.98 ug/l
 RT: 1.837 min Scan# 123
 Delta R.T. -0.049 min
 Lab File: VX022161.D
 Acq: 24 May 2021 20:32




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-938-940MS


Tgt Ion:101 Resp: 79631
 Ion Ratio Lower Upper
 101 100
 103 63.6 53.2 79.8



#8
 Diethyl Ether
 Concen: 52.93 ug/l
 RT: 2.142 min Scan# 173
 Delta R.T. 0.006 min
 Lab File: VX022161.D
 Acq: 24 May 2021 20:32

Tgt Ion: 74 Resp: 34219
 Ion Ratio Lower Upper
 74 100
 45 100.2 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.27 ug/l
RT: 2.300 min Scan# 199
Delta R.T. -0.031 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:101 Resp: 43977
Ion Ratio Lower Upper
101 100
85 47.2 34.9 52.3
151 65.7 61.4 92.2



#10
Methyl Iodide
Concen: 31.84 ug/l
RT: 2.428 min Scan# 220
Delta R.T. -0.025 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:142 Resp: 30545
Ion Ratio Lower Upper
142 100
127 45.8 33.3 49.9
141 13.8 11.4 17.0









#11

Tert butyl alcohol
Concen: 18.77 ug/l
RT: 2.947 min Scan# 305
Delta R.T. -0.042 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS




Tgt Ion: 59 Resp: 4988
Ion Ratio Lower Upper
59 100
57 0.0 7.8 11.8#




#12

1,1-Dichloroethene
Concen: 49.65 ug/l
RT: 2.294 min Scan# 198
Delta R.T. -0.025 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 96 Resp: 42871
Ion Ratio Lower Upper
96 100
61 178.4 124.6 187.0
98 65.3 50.6 76.0









#13
Acrolein
Concen: 245.96 ug/l
RT: 2.252 min Scan# 191
Delta R.T. 0.012 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS

Tgt Ion: 56 Resp: 43362
Ion Ratio Lower Upper
56 100
55 71.1 56.4 84.6



#14
Allyl chloride
Concen: 50.85 ug/l
RT: 2.648 min Scan# 256
Delta R.T. -0.018 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 41 Resp: 80646
Ion Ratio Lower Upper
41 100
39 66.6 51.6 77.4
76 34.4 30.5 45.7









#15
Acrylonitrile
Concen: 303.42 ug/l
RT: 3.093 min Scan# 329
Delta R.T. 0.024 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:	53	Resp:	158781
Ion	Ratio	Lower	Upper
53	100		
52	81.6	65.3	97.9
51	35.2	28.2	42.2



#16
Acetone
Concen: 278.16 ug/l
RT: 2.416 min Scan# 218
Delta R.T. 0.024 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:	43	Resp:	140759
Ion	Ratio	Lower	Upper
43	100		
58	33.4	27.8	41.8









#17
Carbon Disulfide
Concen: 46.87 ug/l
RT: 2.483 min Scan# 229
Delta R.T. -0.031 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 76 Resp: 105991
Ion Ratio Lower Upper
76 100
78 9.5 7.2 10.8



#18
Methyl Acetate
Concen: 61.48 ug/l
RT: 2.727 min Scan# 269
Delta R.T. 0.018 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 43 Resp: 82736
Ion Ratio Lower Upper
43 100
74 23.5 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 56.24 ug/l
RT: 3.136 min Scan# 336
Delta R.T. 0.019 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 73 Resp: 166361
Ion Ratio Lower Upper
73 100
57 21.6 17.1 25.7



#20
Methylene Chloride
Concen: 51.25 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.012 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 84 Resp: 57584
Ion Ratio Lower Upper
84 100
49 119.8 88.6 132.8
51 37.7 27.0 40.6
86 62.9 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 54.15 ug/l
RT: 3.081 min Scan# 327
Delta R.T. -0.018 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 96 Resp: 50858
Ion Ratio Lower Upper
96 100
61 135.2 107.7 161.5
98 65.5 50.4 75.6



#22
Diisopropyl ether
Concen: 56.05 ug/l
RT: 3.788 min Scan# 443
Delta R.T. 0.018 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 45 Resp: 171218
Ion Ratio Lower Upper
45 100
43 54.5 58.0 87.0#
87 26.1 25.4 38.0
59 12.5 10.2 15.2









#23
 Vinyl Acetate
 Concen: 280.48 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. 0.006 min
 Lab File: VX022161.D
 Acq: 24 May 2021 20:32




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-938-940MS


Tgt Ion: 43 Resp: 824906
 Ion Ratio Lower Upper
 43 100
 86 11.1 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 55.81 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.012 min
 Lab File: VX022161.D
 Acq: 24 May 2021 20:32

Tgt Ion: 63 Resp: 100046
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.8 11.2
 100 4.4 2.5 7.5









#25
2-Butanone
Concen: 294.64 ug/l
RT: 4.599 min Scan# 576
Delta R.T. 0.031 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 43 Resp: 241218
Ion Ratio Lower Upper
43 100
72 26.1 23.7 35.5



#26
2,2-Dichloropropane
Concen: 42.80 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 77 Resp: 68112
Ion Ratio Lower Upper
77 100
97 23.0 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 54.86 ug/l

RT: 4.489 min Scan# 558





Delta R.T. -0.006 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 96 Resp: 62656

Ion	Ratio	Lower	Upper
96	100		
61	139.4	0.0	285.6
98	63.5	0.0	129.6



#28

Bromochloromethane

Concen: 49.74 ug/l

RT: 4.910 min Scan# 627




Delta R.T. -0.000 min


Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 49 Resp: 43253

Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	4.0
130	71.7	68.2	102.2





#29

Tetrahydrofuran

Concen: 290.95 ug/l

RT: 5.038 min Scan# 648

Delta R.T. 0.025 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32





Tgt Ion: 42 Resp: 141770

Ion Ratio Lower Upper

42 100

72 45.6 41.7 62.5

71 41.7 38.4 57.6



#30

Chloroform

Concen: 54.27 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX022161.D




Acq: 24 May 2021 20:32

Tgt Ion: 83 Resp: 100508

Ion Ratio Lower Upper

83 100

85 61.5 52.0 78.0




Instrument :

MSVOA_X


ClientSampleId :

BP-VPB-RW8-GW-938-940MS






#31
Cyclohexane
Concen: 51.17 ug/l
RT: 5.465 min Scan# 718
Delta R.T. -0.018 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS



Tgt Ion: 56 Resp: 74734




Ion	Ratio	Lower	Upper
56	100		
69	29.2	25.3	37.9
84	86.5	71.4	107.0



#32
1,1,1-Trichloroethane
Concen: 53.74 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 97 Resp: 85052

Ion	Ratio	Lower	Upper
97	100		
99	64.7	51.5	77.3
61	44.8	34.2	51.4





#33

1,2-Dichloroethane-d4

Concen: 53.15 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min





Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 65 Resp: 61773

Ion Ratio Lower Upper

65	100		
67	53.2	0.0	105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min




Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 114 Resp: 153323

Ion Ratio Lower Upper

114	100		
63	19.2	0.0	40.8
88	16.6	0.0	32.8



Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-RW8-GW-938-940MS



#35
Dibromofluoromethane
Concen: 50.77 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:113 Resp: 49368
Ion Ratio Lower Upper
113 100
111 101.3 81.9 122.9
192 16.4 16.7 25.1#



#36
1,1-Dichloropropene
Concen: 50.87 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.012 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32


Tgt Ion: 75 Resp: 72121
Ion Ratio Lower Upper
75 100
110 34.8 18.1 54.4
77 30.7 25.3 37.9








#37
Ethyl Acetate
Concen: 54.07 ug/l
RT: 4.745 min Scan# 600
Delta R.T. 0.024 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS







Tgt Ion: 43 Resp: 93968
Ion Ratio Lower Upper
43 100
61 14.5 12.8 19.2
70 10.9 10.5 15.7



#38
Carbon Tetrachloride
Concen: 50.49 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.018 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 117 Resp: 70023
Ion Ratio Lower Upper
117 100
119 92.6 76.2 114.2
121 33.5 24.8 37.2





#39

Methylcyclohexane

Concen: 47.23 ug/l

RT: 7.379 min Scan# 1032





Delta R.T. -0.006 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 83 Resp: 75946

Ion	Ratio	Lower	Upper
83	100		
55	73.1	58.9	88.3
98	50.9	38.1	57.1



#40

Benzene

Concen: 51.85 ug/l

RT: 6.038 min Scan# 812




Delta R.T. -0.006 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 78 Resp: 226176


Ion	Ratio	Lower	Upper
78	100		
77	22.2	18.9	28.3








#41
Methacrylonitrile
Concen: 54.28 ug/l
RT: 4.946 min Scan# 633
Delta R.T. 0.018 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS







Tgt Ion:	41	Resp:	50472
Ion	Ratio	Lower	Upper
41	100		
39	53.7	42.5	63.7
67	73.7	65.0	97.6
52	29.3	25.3	37.9



#42
1,2-Dichloroethane
Concen: 52.31 ug/l
RT: 6.099 min Scan# 822
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:	62	Resp:	79940
Ion	Ratio	Lower	Upper
62	100		
98	9.4	0.0	20.2









#43
Isopropyl Acetate
Concen: 55.08 ug/l
RT: 6.361 min Scan# 865
Delta R.T. 0.012 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS

Tgt Ion: 43 Resp: 156575
Ion Ratio Lower Upper
43 100
61 19.8 18.2 27.4
87 13.6 12.5 18.7



#44
Trichloroethene
Concen: 49.62 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:130 Resp: 53791
Ion Ratio Lower Upper
130 100
95 99.1 0.0 198.0





#45

1,2-Dichloropropane

Concen: 54.47 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 63 Resp: 60610

Ion Ratio Lower Upper

63 100





65 30.9 25.7 38.5

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-RW8-GW-938-940MS



#46

Dibromomethane

Concen: 52.80 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32




Tgt Ion: 93 Resp: 39864

Ion Ratio Lower Upper

93 100

95 82.3 66.8 100.2

174 80.5 83.1 124.7#









#47
Bromodichloromethane
Concen: 51.99 ug/l
RT: 7.830 min Scan# 1106
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS

Tgt Ion: 83 Resp: 78542
Ion Ratio Lower Upper
83 100
85 63.8 51.4 77.0
127 7.4 6.8 10.2



#48
Methyl methacrylate
Concen: 54.15 ug/l
RT: 7.702 min Scan# 1085
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 41 Resp: 70945
Ion Ratio Lower Upper
41 100
69 86.8 76.0 114.0
39 59.4 44.6 67.0





#49

1,4-Dioxane

Concen: 1089.29 ug/l

RT: 7.690 min Scan# 1083

Delta R.T. 0.006 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32





Tgt Ion: 88 Resp: 27785

Ion Ratio Lower Upper

88 100

43 32.4 23.2 34.8

58 68.6 55.0 82.4



#50

Toluene-d8

Concen: 50.18 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX022161.D




Acq: 24 May 2021 20:32

Tgt Ion: 98 Resp: 187951

Ion Ratio Lower Upper

98 100

100 65.3 51.9 77.9




Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-RW8-GW-938-940MS



#51
4-Methyl-2-Pentanone
Concen: 292.15 ug/l
RT: 8.592 min Scan# 1231
Delta R.T. 0.012 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS

Tgt Ion: 43 Resp: 497485
Ion Ratio Lower Upper
43 100
58 39.5 35.4 53.0



#52
Toluene
Concen: 52.05 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 92 Resp: 143788
Ion Ratio Lower Upper
92 100
91 166.5 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 50.13 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 75 Resp: 87418
Ion Ratio Lower Upper
75 100
77 32.9 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 50.86 ug/l
RT: 8.373 min Scan# 1195
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 75 Resp: 95344
Ion Ratio Lower Upper
75 100
77 30.8 25.7 38.5
39 48.1 35.1 52.7





#55

1,1,2-Trichloroethane

Concen: 53.60 ug/l

RT: 9.159 min Scan# 1324

Delta R.T. -0.000 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32

Tgt Ion: 97 Resp: 60112





Ion Ratio Lower Upper

97 100

83 86.0 69.3 103.9

85 57.3 44.3 66.5

99 62.1 50.3 75.5



#56

Ethyl methacrylate

Concen: 54.86 ug/l

RT: 9.128 min Scan# 1319

Delta R.T. 0.006 min

Lab File: VX022161.D

Acq: 24 May 2021 20:32




Tgt Ion: 69 Resp: 99427


Ion Ratio Lower Upper

69 100

41 64.5 46.8 70.2

39 38.5 24.7 37.1#









#57

1,3-Dichloropropane
Concen: 53.97 ug/l
RT: 9.317 min Scan# 1350
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS




Tgt Ion: 76 Resp: 103342
Ion Ratio Lower Upper
76 100
78 32.3 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 258.27 ug/l
RT: 8.251 min Scan# 1175
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 63 Resp: 253183
Ion Ratio Lower Upper
63 100
106 27.8 23.5 35.3









#59
2-Hexanone
Concen: 301.55 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 43 Resp: 386199
Ion Ratio Lower Upper
43 100
58 55.5 30.9 92.8



#60
Dibromochloromethane
Concen: 53.09 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:129 Resp: 57920
Ion Ratio Lower Upper
129 100
127 78.8 38.4 115.2








#61
1,2-Dibromoethane
Concen: 54.56 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 107 Resp: 60917
Ion Ratio Lower Upper
107 100
109 96.0 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 54.10 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 95 Resp: 74891
Ion Ratio Lower Upper
95 100
174 64.3 0.0 154.6
176 61.7 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.061 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:117 Resp: 140860
Ion Ratio Lower Upper
117 100
82 56.7 47.4 71.2
119 32.8 26.6 39.8



#64
Tetrachloroethene
Concen: 48.04 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32


Tgt Ion:164 Resp: 44426
Ion Ratio Lower Upper
164 100
166 128.5 103.4 155.2
129 105.6 70.8 106.2
131 99.1 72.5 108.7








#65
Chlorobenzene
Concen: 51.07 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32


Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS





Tgt Ion:112 Resp: 149763
Ion Ratio Lower Upper
112 100
114 32.1 25.8 38.6




#66
1,1,1,2-Tetrachloroethane
Concen: 52.38 ug/l
RT: 10.171 min Scan# 1490
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32



Tgt Ion:131 Resp: 53375
Ion Ratio Lower Upper
131 100
133 97.4 47.7 143.1
119 67.3 32.1 96.5









#67
Ethyl Benzene
Concen: 51.32 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 91 Resp: 276828
Ion Ratio Lower Upper
91 100
106 30.6 24.6 37.0



#68
m/p-Xylenes
Concen: 103.65 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 106 Resp: 211315
Ion Ratio Lower Upper
106 100
91 204.0 162.8 244.2









#69
o-Xylene
Concen: 51.94 ug/l
RT: 10.647 min Scan# 1568
Delta R.T. -0.001 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS

Tgt Ion:106 Resp: 102983
Ion Ratio Lower Upper
106 100
91 214.6 108.1 324.2



#70
Styrene
Concen: 54.25 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32


Tgt Ion:104 Resp: 176203
Ion Ratio Lower Upper
104 100
78 52.2 40.0 60.0
103 54.2 43.4 65.0








#71
Bromoform
Concen: 50.22 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32


Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS





Tgt Ion:173 Resp: 39218
Ion Ratio Lower Upper
173 100
175 46.6 23.9 71.8
254 0.0 0.1 0.1#




#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32



Tgt Ion:152 Resp: 68489
Ion Ratio Lower Upper
152 100
115 91.6 41.1 123.3
150 174.9 0.0 349.0









#73
Isopropylbenzene
Concen: 46.63 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:105 Resp: 270291
Ion Ratio Lower Upper
105 100
120 26.3 13.2 39.6



#74
N-ethyl acetate
Concen: 54.25 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 43 Resp: 149271
Ion Ratio Lower Upper
43 100
70 42.7 39.0 58.6
55 23.0 21.5 32.3
61 24.6 22.1 33.1









#75

1,1,2,2-Tetrachloroethane
Concen: 54.99 ug/l
RT: 11.220 min Scan# 1662
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS




Tgt Ion: 83 Resp: 105660
Ion Ratio Lower Upper
83 100
131 9.1 5.1 15.2
85 63.8 31.8 95.4




#76

1,2,3-Trichloropropane
Concen: 69.36 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 75 Resp: 117189
Ion Ratio Lower Upper
75 100
77 33.0 19.1 57.5









#77
Bromobenzene
Concen: 48.96 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:156 Resp: 60173
Ion Ratio Lower Upper
156 100
77 186.9 82.8 248.6
158 96.7 49.1 147.4



#78
n-propylbenzene
Concen: 47.88 ug/l
RT: 11.311 min Scan# 1677
Delta R.T. 0.006 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 91 Resp: 325095
Ion Ratio Lower Upper
91 100
120 22.2 11.5 34.5









#79
2-Chlorotoluene
Concen: 46.94 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 91 Resp: 193298
Ion Ratio Lower Upper
91 100
126 32.3 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 49.29 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 105 Resp: 236827
Ion Ratio Lower Upper
105 100
120 47.4 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 47.84 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. -0.001 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 75 Resp: 34114
Ion Ratio Lower Upper
75 100
53 89.4 74.5 111.7
89 43.0 37.9 56.9



#82
4-Chlorotoluene
Concen: 47.73 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 91 Resp: 226475
Ion Ratio Lower Upper
91 100
126 28.2 14.9 44.5









#83
tert-Butylbenzene
Concen: 49.32 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. -0.001 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:119 Resp: 227264
Ion Ratio Lower Upper
119 100
91 62.4 27.9 83.7
134 22.8 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 50.21 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:105 Resp: 242826
Ion Ratio Lower Upper
105 100
120 44.4 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 48.98 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX022161.D
 Acq: 24 May 2021 20:32




Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-GW-938-940MS


Tgt Ion:105 Resp: 286791
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 49.27 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX022161.D
 Acq: 24 May 2021 20:32

Tgt Ion:119 Resp: 243653
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.9 38.7
 91 25.9 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 49.69 ug/l
RT: 11.976 min Scan# 1786
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:146 Resp: 117753
Ion Ratio Lower Upper
146 100
111 45.1 20.2 60.5
148 62.3 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 49.66 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:146 Resp: 121636
Ion Ratio Lower Upper
146 100
111 44.0 19.6 58.8
148 61.8 31.6 95.0









#89
n-Butylbenzene
Concen: 50.51 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion: 91 Resp: 231043
Ion Ratio Lower Upper
91 100
92 53.5 27.4 82.0
134 24.2 13.2 39.5



#90
Hexachloroethane
Concen: 48.89 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 117 Resp: 39962
Ion Ratio Lower Upper
117 100
201 59.9 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 51.85 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:146 Resp: 118028
Ion Ratio Lower Upper
146 100
111 47.0 20.8 62.2
148 64.0 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 56.85 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion: 75 Resp: 23372
Ion Ratio Lower Upper
75 100
155 75.7 45.3 135.9
157 95.4 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 52.49 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS


Tgt Ion:180 Resp: 72092
Ion Ratio Lower Upper
180 100
182 96.3 47.7 143.1
145 35.8 15.0 45.0



#94
Hexachlorobutadiene
Concen: 50.49 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:225 Resp: 27070
Ion Ratio Lower Upper
225 100
223 61.4 31.1 93.2
227 63.2 31.6 95.0









#95
Naphthalene
Concen: 56.26 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-GW-938-940MS

Tgt Ion:128 Resp: 301381
Ion Ratio Lower Upper
128 100
127 12.7 10.3 15.5
129 10.6 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 52.78 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX022161.D
Acq: 24 May 2021 20:32

Tgt Ion:180 Resp: 73076
Ion Ratio Lower Upper
180 100
182 95.3 47.6 142.9
145 37.0 16.1 48.3

