

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052421\
 Data File : VX022143.D
 Acq On : 24 May 2021 13:22
 Operator : JC/MD
 Sample : VX0524WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0524WBL01

Quant Time: May 24 15:51:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	71901	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	149104	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49831	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57185	51.56	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.12%
35) Dibromofluoromethane	5.391	113	45618	48.25	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.50%
50) Toluene-d8	8.653	98	178314	48.96	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.92%
62) 4-Bromofluorobenzene	11.085	95	61811	45.91	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052421\
Data File : VX022143.D
Acq On : 24 May 2021 13:22
Operator : JC/MD
Sample : VX0524WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0524WBL01

Quant Time: May 24 15:51:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

