

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052421\
 Data File : VX022149.D
 Acq On : 24 May 2021 15:48
 Operator : JC/MD
 Sample : M2498-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW8-TB-20210520

Quant Time: May 25 02:07:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	70127	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	142444	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	121144	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48373	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	54599	50.47	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.94%	
35) Dibromofluoromethane	5.397	113	44463	49.22	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.44%	
50) Toluene-d8	8.653	98	170841	49.10	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.20%	
62) 4-Bromofluorobenzene	11.085	95	58754	45.68	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.36%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052421\
Data File : VX022149.D
Acq On : 24 May 2021 15:48
Operator : JC/MD
Sample : M2498-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW8-TB-20210520

Quant Time: May 25 02:07:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

