

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
 Data File : VX028921.D
 Acq On : 24 May 2022 15:08
 Operator : JC/MD
 Sample : N2943-17DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-38IDL

Quant Time: May 25 01:49:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	341411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	603704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	578867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	273400	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	233194	49.828	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.660%
35) Dibromofluoromethane	5.391	113	199895	50.537	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.080%
50) Toluene-d8	8.653	98	712762	47.796	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.600%
62) 4-Bromofluorobenzene	11.079	95	288648	51.823	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.640%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	458427	36.560	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
Data File : VX028921.D
Acq On : 24 May 2022 15:08
Operator : JC/MD
Sample : N2943-17DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-38IDL

Quant Time: May 25 01:49:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

