

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
 Data File : VX028922.D
 Acq On : 24 May 2022 15:32
 Operator : JC/MD
 Sample : N2970-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 70BR-72-20220519DL

Quant Time: May 25 01:49:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	596114	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	570538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	231764	49.641	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.280%
35) Dibromofluoromethane	5.391	113	197049	50.452	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.900%
50) Toluene-d8	8.653	98	706988	48.013	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.020%
62) 4-Bromofluorobenzene	11.079	95	279852	50.883	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.760%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	9045	1.998	ug/l	99
21) trans-1,2-Dichloroethene	3.093	96	1421	0.358	ug/l #	76
27) cis-1,2-Dichloroethene	4.489	96	159292	34.349	ug/l	88
44) Trichloroethene	7.129	130	153430	33.556	ug/l	98
52) Toluene	8.720	92	2704	0.249	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
Data File : VX028922.D
Acq On : 24 May 2022 15:32
Operator : JC/MD
Sample : N2970-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
70BR-72-20220519DL

Quant Time: May 25 01:49:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

