

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
 Data File : VX028923.D
 Acq On : 24 May 2022 15:55
 Operator : JC/MD
 Sample : N2965-07DL 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31608DL

Quant Time: May 24 17:22:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	330622	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	590798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	571159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	275402	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	231678	51.120	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.240%	
35) Dibromofluoromethane	5.391	113	194277	50.190	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.380%	
50) Toluene-d8	8.653	98	710420	48.680	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.360%	
62) 4-Bromofluorobenzene	11.079	95	286490	52.559	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.120%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1739	0.894	ug/l	95
31) Cyclohexane	5.477	56	16986	2.700	ug/l	96
39) Methylcyclohexane	7.385	83	7679	1.152	ug/l	# 86
40) Benzene	6.044	78	524706	31.402	ug/l	100
52) Toluene	8.720	92	1354664	126.032	ug/l	96
67) Ethyl Benzene	10.195	91	217614	10.020	ug/l	100
68) m/p-Xylenes	10.299	106	320064	37.853	ug/l	97
69) o-Xylene	10.640	106	149244	17.928	ug/l	96
73) Isopropylbenzene	10.963	105	7697	0.369	ug/l	98
78) n-propylbenzene	11.305	91	21249	0.880	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	28826	1.656	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	124856	7.155	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
Data File : VX028923.D
Acq On : 24 May 2022 15:55
Operator : JC/MD
Sample : N2965-07DL 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31608DL

Quant Time: May 24 17:22:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

