

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
 Data File : VX028925.D
 Acq On : 24 May 2022 16:42
 Operator : JC/MD
 Sample : N2970-08DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 70BR-10-20220519DL

Quant Time: May 25 01:49:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	347810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	617121	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	598225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	287255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	236579	49.622	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.240%		
35) Dibromofluoromethane	5.385	113	201322	49.792	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.580%		
50) Toluene-d8	8.653	98	737456	48.377	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.760%		
62) 4-Bromofluorobenzene	11.079	95	298142	52.363	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	104.720%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	44092	9.311	ug/l	90
44) Trichloroethene	7.129	130	278995	58.940	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
Data File : VX028925.D
Acq On : 24 May 2022 16:42
Operator : JC/MD
Sample : N2970-08DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
70BR-10-20220519DL

Quant Time: May 25 01:49:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

