

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
 Data File : VX028927.D
 Acq On : 24 May 2022 17:28
 Operator : JC/MD
 Sample : N2882-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SLP-2

Quant Time: May 25 01:50:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	337638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	595608	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	579236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	276288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	228742	49.423	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.840%
35) Dibromofluoromethane	5.385	113	197209	50.536	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.080%
50) Toluene-d8	8.653	98	712503	48.428	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.860%
62) 4-Bromofluorobenzene	11.079	95	286753	52.182	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
Data File : VX028927.D
Acq On : 24 May 2022 17:28
Operator : JC/MD
Sample : N2882-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SLP-2

Quant Time: May 25 01:50:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

