

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
 Data File : VX028931.D
 Acq On : 24 May 2022 19:02
 Operator : JC/MD
 Sample : N2954-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB-220518

Quant Time: May 25 01:51:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	321885	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	575791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	549933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	254518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	223146	50.574	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.140%			
35) Dibromofluoromethane	5.391	113	190829	50.584	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.160%			
50) Toluene-d8	8.653	98	683036	48.023	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.040%			
62) 4-Bromofluorobenzene	11.079	95	269567	50.743	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.480%			
Target Compounds						
16) Acetone	2.380	43	6643	3.508	ug/l	92
29) Tetrahydrofuran	5.025	42	8836	4.573	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
Data File : VX028931.D
Acq On : 24 May 2022 19:02
Operator : JC/MD
Sample : N2954-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB-220518

Quant Time: May 25 01:51:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

