

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
 Data File : VX028935.D
 Acq On : 24 May 2022 20:35
 Operator : JC/MD
 Sample : N2980-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: May 25 01:52:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	326792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	581244	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	551178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	255461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	228318	50.969	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.940%	
35) Dibromofluoromethane	5.391	113	196888	51.701	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.400%	
50) Toluene-d8	8.653	98	689240	48.005	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.000%	
62) 4-Bromofluorobenzene	11.079	95	272909	50.890	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.780%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4410	2.294	ug/l	95
20) Methylene Chloride	2.788	84	2247	0.557	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052422\
Data File : VX028935.D
Acq On : 24 May 2022 20:35
Operator : JC/MD
Sample : N2980-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: May 25 01:52:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

