

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
 Data File : VX035831.D
 Acq On : 24 May 2023 13:15
 Operator : JC/MD
 Sample : 02860-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-DUP-03-05162023

Quant Time: May 25 05:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	188439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	332997	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	291235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139855	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	140335	42.689	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.380%
35) Dibromofluoromethane	5.385	113	107337	47.796	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.600%
50) Toluene-d8	8.646	98	364531	43.939	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.880%#
62) 4-Bromofluorobenzene	11.079	95	145636	43.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.200%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	3127	2.213	ug/l	93
17) Carbon Disulfide	2.507	76	1706	0.337	ug/l	98
44) Trichloroethene	7.122	130	327351	130.187	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
Data File : VX035831.D
Acq On : 24 May 2023 13:15
Operator : JC/MD
Sample : 02860-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-DUP-03-05162023

Quant Time: May 25 05:30:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

