

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
 Data File : VX035836.D
 Acq On : 24 May 2023 15:15
 Operator : JC/MD
 Sample : 02911-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20230522

Quant Time: May 25 05:32:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	174871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	302168	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	269610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127996	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	124434	40.789	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.580%
35) Dibromofluoromethane	5.385	113	95958	47.089	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.180%
50) Toluene-d8	8.647	98	338652	44.985	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.960%#
62) 4-Bromofluorobenzene	11.079	95	134606	44.407	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.820%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	2099	1.601	ug/l	94
19) Methyl tert-butyl Ether	3.111	73	3477	0.449	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	4261	1.661	ug/l	84
30) Chloroform	5.093	83	909	0.203	ug/l	90
44) Trichloroethene	7.129	130	1853	0.812	ug/l	83
64) Tetrachloroethene	9.275	164	13422	7.395	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
Data File : VX035836.D
Acq On : 24 May 2023 15:15
Operator : JC/MD
Sample : 02911-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20230522

Quant Time: May 25 05:32:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

