

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
 Data File : VX035839.D
 Acq On : 24 May 2023 16:24
 Operator : JC/MD
 Sample : 02911-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB02-20230522

Quant Time: May 25 05:33:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	170858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	300545	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	126313	42.377	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.760%
35) Dibromofluoromethane	5.385	113	95582	47.158	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.320%
50) Toluene-d8	8.647	98	329035	43.943	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.880%#
62) 4-Bromofluorobenzene	11.079	95	129780	43.046	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
Data File : VX035839.D
Acq On : 24 May 2023 16:24
Operator : JC/MD
Sample : 02911-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB02-20230522

Quant Time: May 25 05:33:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

