

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
 Data File : VX035843.D
 Acq On : 24 May 2023 17:58
 Operator : JC/MD
 Sample : 02911-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D2-20230522

Quant Time: May 25 05:34:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	288807	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	122624	42.512	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	85.020%		
35) Dibromofluoromethane	5.385	113	92143	47.308	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.620%		
50) Toluene-d8	8.647	98	315476	43.845	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	87.680%#		
62) 4-Bromofluorobenzene	11.079	95	124949	43.128	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.260%		
Target Compounds						Qvalue
16) Acetone	2.404	43	1851	1.493	ug/l #	74
29) Tetrahydrofuran	5.025	42	18052	16.735	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052423\
Data File : VX035843.D
Acq On : 24 May 2023 17:58
Operator : JC/MD
Sample : 02911-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D2-20230522

Quant Time: May 25 05:34:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

